



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: JIGNESH

Run Number: LB126385

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	07/11/2023	08:20
2	O3549-01		1	12	100.00	>212.0	>212.0	07/11/2023	08:50
3	O3549-01DUP		1	14	100.00	>212.0	>212.0	07/11/2023	09:20
4	O3549-02		1	13	100.00	>212.0	>212.0	07/11/2023	09:50
5	O3562-06		1	12	100.00	>212.0	>212.0	07/11/2023	10:20
6	O3563-01		1	14	100.00	>212.0	>212.0	07/11/2023	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126385

WorkList Name : fp 7-10

WorkList ID : 171758

Department : Wet-Chemistry

Date : 07-10-2023 13:27:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3549-01 C	60101	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/07/2023	1010B
O3549-02 E	60102	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/07/2023	1010B
O3562-06 D	34590	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/07/2023	1010B
O3563-01 J	HOUSE-BASEMENT	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/07/2023	1010B

Date/Time 07/11/2023 08:10
 Raw Sample Received by: Roy Cw12
 Raw Sample Relinquished by: J. L. Cw12

Date/Time 07/11/2023 11:30
 Raw Sample Received by: J. L. Cw12
 Raw Sample Relinquished by: Roy Cw12