



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126392

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	07/12/2023	07:50
2	CAL2	1	Water	NA	NA	20.5	7.00	07/12/2023	07:51
3	CAL3	1	Water	NA	NA	20.7	10.02	07/12/2023	07:53
4	ICV	1	Water	NA	NA	20.7	6.99	07/12/2023	08:00
5	CCV1	1	Water	NA	NA	20.6	2.01	07/12/2023	08:02
6	O3574-01	1	Water	NA	NA	22.2	4.81	07/12/2023	08:10
7	O3574-01DUP	1	Water	NA	NA	22.3	4.82	07/12/2023	08:11
8	CCV2	1	Water	NA	NA	20.7	12.02	07/12/2023	08:15

126392

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o3474

WorkList ID : 171785

Department : Wet-Chemistry

Date : 07-12-2023 07:43:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3574-01	A	POOL-WATER	Water	pH	STER11	H21	07/06/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: