

LB126404


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46192



A= 0.0000e+000

B= 2.1478e-004

C= -4.4059e-002

Rho= 0.9997401

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	7/0/0
0.0	0.000	0.006	0.006	231	0.000	231	0				-
0.2	0.200	0.179	-0.021	1039	0.0 %	1039					-11
2.5	2.500	2.454	-0.046	11633	0.0 %	11633					-2
5.0	5.000	5.029	0.029	23622	0.0 %	23622					1
7.5	7.500	7.655	0.155	35845	0.0 %	35845					2
10.0	10.000	9.877	-0.123	46192	0.0 %	46192					-1

LB126404 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jul 2023 13:22:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	231	-	0.0000	7470A	PPB	12 Jul 2023 13:25:45	S	Std
0.2 - 1	50.2	1039	-	0.2000	7470A	PPB	12 Jul 2023 13:28:02	S	Std
2.5 - 1	52.5	11633	-	2.5000	7470A	PPB	12 Jul 2023 13:30:19	S	Std
5.0 - 1	55	23622	-	5.0000	7470A	PPB	12 Jul 2023 13:32:36	S	Std
7.5 - 1	57.5	35845	-	7.5000	7470A	PPB	12 Jul 2023 13:34:52	S	Std
10.0 - 1	510	46192	-	10.0000	7470A	PPB	12 Jul 2023 13:37:07	S	Std
ICV65 - 1	ICV65	18006	3.8232	-	7470A	PPB	12 Jul 2023 13:42:33	U	SMPL
ICB65 - 1	ICB65	211	0.0013	-	7470A	PPB	12 Jul 2023 13:44:49	U	SMPL
CCV18 - 1	CCV18	23194	4.9375	-	7470A	PPB	12 Jul 2023 13:47:05	U	SMPL
CCB18 - 1	CCB18	229	0.0051	-	7470A	PPB	12 Jul 2023 13:49:19	U	SMPL
CRA - 1	CRA	1025	0.1761	-	7470A	PPB	12 Jul 2023 13:51:38	U	SMPL
HighStd - 1	HighStd	45163	9.6559	-	7470A	PPB	12 Jul 2023 13:53:53	U	SMPL
ChkStd - 1	ChkStd	34415	7.3475	-	7470A	PPB	12 Jul 2023 13:56:08	U	SMPL
PB154106BL - 1	PBW	50	-0.0333	-	7470A	PPB	12 Jul 2023 13:58:24	U	SMPL
PB154106BS - 1	LCSW	17440	3.7016	-	7470A	PPB	12 Jul 2023 14:00:41	U	SMPL
O3541-02 - 1	WC-WWS-08	80	-0.0269	-	7470A	PPB	12 Jul 2023 14:02:57	U	SMPL
O3541-07 - 1	WC-WWS-09	172	-0.0071	-	7470A	PPB	12 Jul 2023 14:05:13	U	SMPL
O3555-02 - 1	ETGI-330	214	0.0019	-	7470A	PPB	12 Jul 2023 14:07:30	U	SMPL
O3555-04 - 1	ETGI-347	186	-0.0041	-	7470A	PPB	12 Jul 2023 14:09:46	U	SMPL
O3555-06 - 1	ETGI-369	188	-0.0037	-	7470A	PPB	12 Jul 2023 14:12:02	U	SMPL
CCV19 - 1	CCV19	23406	4.9830	-	7470A	PPB	12 Jul 2023 14:14:17	U	SMPL
CCB19 - 1	CCB19	241	0.0077	-	7470A	PPB	12 Jul 2023 14:16:32	U	SMPL
O3555-08 - 1	VNJ-203	178	-0.0058	-	7470A	PPB	12 Jul 2023 14:18:47	U	SMPL
O3555-10 - 1	ETGI-289	297	0.0197	-	7470A	PPB	12 Jul 2023 14:21:03	U	SMPL
O3555-12 - 1	VNJ-240	264	0.0126	-	7470A	PPB	12 Jul 2023 14:23:18	U	SMPL
O3555-12DUP - 1	VNJ-240DUP	268	0.0135	-	7470A	PPB	12 Jul 2023 14:25:33	U	SMPL
O3555-12MS - 1	VNJ-240MS	20226	4.3000	-	7470A	PPB	12 Jul 2023 14:27:49	U	SMPL
O3555-12MSD - 1	VNJ-240MSD	19972	4.2455	-	7470A	PPB	12 Jul 2023 14:30:05	U	SMPL
PB154115BL - 1	PBW	80	-0.0269	-	7470A	PPB	12 Jul 2023 14:32:22	U	SMPL
PB154115BS - 1	LCSW	18307	3.8879	-	7470A	PPB	12 Jul 2023 14:34:39	U	SMPL
O3562-06 - 1	34590	930	0.1557	-	7470A	PPB	12 Jul 2023 14:41:59	U	SMPL
O3562-06DUP - 1	34590DUP	885	0.1460	-	7470A	PPB	12 Jul 2023 14:44:14	U	SMPL
CCV20 - 1	CCV20	23754	5.0577	-	7470A	PPB	12 Jul 2023 14:46:30	U	SMPL
CCB20 - 1	CCB20	180	-0.0054	-	7470A	PPB	12 Jul 2023 14:48:45	U	SMPL
O3562-06MS - 1	34590MS	16797	3.5635	-	7470A	PPB	12 Jul 2023 14:51:00	U	SMPL
O3562-06MSD - 1	34590MSD	15284	3.2386	-	7470A	PPB	12 Jul 2023 14:53:16	U	SMPL
O3563-01 - 1	HOUSE-BASEMENT	3652	0.7403	-	7470A	PPB	12 Jul 2023 14:55:32	U	SMPL
PB154034TB - 1	PB154034TB	166	-0.0084	-	7470A	PPB	12 Jul 2023 14:57:48	U	SMPL
O3555-12LX5 - 1		38	-0.0359	-	7470A	PPB	12 Jul 2023 15:00:03	U	SMPL
O3555-12A - 1		19260	4.0925	-	7470A	PPB	12 Jul 2023 15:02:20	U	SMPL
O3562-06LX5 - 1		134	-0.0153	-	7470A	PPB	12 Jul 2023 15:04:36	U	SMPL
O3562-06A - 1		21254	4.5208	-	7470A	PPB	12 Jul 2023 15:06:53	U	SMPL
CCV21 - 1	CCV21	23885	5.0859	-	7470A	PPB	12 Jul 2023 15:09:10	U	SMPL
CCB21 - 1	CCB21	-83	-0.0619	-	7470A	PPB	12 Jul 2023 15:11:27	U	SMPL