


LB126408

245.1

INSTRUMENT ID: CV3



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.005	0.005	5	0.000	5					1
0.05	0.050	0.049	-0.001	193	0.0 %	193	0				-2
0.2	0.200	0.245	0.045	1034	0.0 %	1034					23
2.5	2.500	2.504	0.004	10701	0.0 %	10701					0
5.0	5.000	5.042	0.042	21561	0.0 %	21561					1
7.5	7.500	7.249	-0.251	31002	0.0 %	31002					-3
10.0	10.000	10.156	0.156	43444	0.0 %	43444					2

LB126408 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Jul 2023 15:29:41

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	245.1	PPB	12 Jul 2023 15:39:05	S	Std
0.05 - 1	50.45	193	-	0.0500	245.1	PPB	12 Jul 2023 15:41:22	S	Std
0.2 - 1	51.2	1034	-	0.2000	245.1	PPB	12 Jul 2023 15:43:39	S	Std
2.5 - 1	52.5	10701	-	2.5000	245.1	PPB	12 Jul 2023 15:45:56	S	Std
5.0 - 1	55	21561	-	5.0000	245.1	PPB	12 Jul 2023 15:50:51	S	Std
7.5 - 1	57.5	31002	-	7.5000	245.1	PPB	12 Jul 2023 15:58:15	S	Std
10.0 - 1	510	43444	-	10.0000	245.1	PPB	12 Jul 2023 16:11:15	S	Std
ICV66 - 1	ICV66	16912	3.9558	-245.1		PPB	12 Jul 2023 16:17:02	U	SMPL
ICB66 - 1	ICB66	-116	-0.0236	-245.1		PPB	12 Jul 2023 16:19:17	U	SMPL
CCV22 - 1	CCV22	21757	5.0881	-245.1		PPB	12 Jul 2023 16:21:36	U	SMPL
CCB22 - 1	CCB22	-74	-0.0138	-245.1		PPB	12 Jul 2023 16:23:51	U	SMPL
CRA - 1	CRA	174	0.0442	-245.1		PPB	12 Jul 2023 16:26:06	U	SMPL
HighStd - 1	HighStd	43907	10.2645	-245.1		PPB	12 Jul 2023 16:28:21	U	SMPL
ChkStd - 1	ChkStd	29028	6.7873	-245.1		PPB	12 Jul 2023 16:30:35	U	SMPL
PB154152BL - 1	PBW	-159	-0.0336	-245.1		PPB	12 Jul 2023 16:32:58	U	SMPL
PB154152BS - 1	LCSW	16045	3.7532	-245.1		PPB	12 Jul 2023 16:35:13	U	SMPL
O3520-01 - 1	402	-265	-0.0584	-245.1		PPB	12 Jul 2023 16:37:30	U	SMPL
O3520-01DUP - 1	402DUP	197	0.0496	-245.1		PPB	12 Jul 2023 16:39:56	U	SMPL
O3520-01MS - 1	402MS	15926	3.7254	-245.1		PPB	12 Jul 2023 16:42:13	U	SMPL
O3520-01MSD - 1	402MSD	16074	3.7600	-245.1		PPB	12 Jul 2023 16:44:30	U	SMPL
CCV23 - 1	CCV23	21002	4.9116	-245.1		PPB	12 Jul 2023 16:46:46	U	SMPL
CCB23 - 1	CCB23	-92	-0.0180	-245.1		PPB	12 Jul 2023 16:49:03	U	SMPL
O3520-01LX5 - 1		56	0.0166	-245.1		PPB	12 Jul 2023 16:51:21	U	SMPL
O3520-01A - 1		15724	3.6782	-245.1		PPB	12 Jul 2023 16:53:36	U	SMPL
CCV24 - 1	CCV24	21405	5.0058	-245.1		PPB	12 Jul 2023 16:55:51	U	SMPL
CCB24 - 1	CCB24	-101	-0.0201	-245.1		PPB	12 Jul 2023 16:58:06	U	SMPL