

LB126412


SFAM01.1

INSTRUMENT ID: C01

Linear

 μ Abs.:

54223



A= 0.0000e+000

B= 1.8305e-004 slope

C= -7.6201e-002 y-intercept

Rho= 0.9996686

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.00	0.000	-0.037	-0.037	214	0.000	214					>
0.05	0.050					302	0				-12
0.20	0.200	0.177	-0.023	1381	0.0 %	1381					0
2.50	2.500	2.508	0.008	14117	0.0 %	14117					1
5.00	5.000	5.035	0.035	27920	0.0 %	27920					2
7.50	7.500	7.668	0.168	42307	0.0 %	42307					-2
10.0	10.000	9.850	-0.150	54223	0.0 %	54223					

LB126412

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	214	-		0 SFAM01.1	PPB	7/12/2023 17:32	Std
	0.2 S01	1381	-		0.2 SFAM01.1	PPB	7/12/2023 17:35	Std
	2.5 S02	14117	-		2.5 SFAM01.1	PPB	7/12/2023 17:39	Std
	5 S03	27920	-		5 SFAM01.1	PPB	7/12/2023 17:41	Std
	7.5 S04	42307	-		7.5 SFAM01.1	PPB	7/12/2023 17:44	Std
	10 S05	54223	-		10 SFAM01.1	PPB	7/12/2023 17:46	Std
ICV067	ICV067	24382	4.387 -		SFAM01.1	PPB	7/12/2023 17:49	SMPL
ICB067	ICB067	174	-0.0443 -		SFAM01.1	PPB	7/12/2023 17:51	SMPL
CCV025	CCV025	28170	5.0804 -		SFAM01.1	PPB	7/12/2023 17:53	SMPL
CCB025	CCB025	172	-0.0447 -		SFAM01.1	PPB	7/12/2023 17:56	SMPL
PB154157BL	PBS157	643	0.0415 -		SFAM01.1	PPB	7/12/2023 17:58	SMPL
O3460-03	A46L3	2504	0.3822 -		SFAM01.1	PPB	7/12/2023 18:00	SMPL
O3460-04	A46L8	1773	0.2484 -		SFAM01.1	PPB	7/12/2023 18:02	SMPL
O3460-05	A46L9	3196	0.5088 -		SFAM01.1	PPB	7/12/2023 18:05	SMPL
O3460-06	A46M0	2341	0.3523 -		SFAM01.1	PPB	7/12/2023 18:07	SMPL
O3460-07	A46M1	2789	0.4343 -		SFAM01.1	PPB	7/12/2023 18:09	SMPL
O3460-08	A46M2	1229	0.1488 -		SFAM01.1	PPB	7/12/2023 18:11	SMPL
O3460-09	A46M3	712	0.0541 -		SFAM01.1	PPB	7/12/2023 18:14	SMPL
O3460-10	A46M5	23988	4.3149 -		SFAM01.1	PPB	7/12/2023 18:16	SMPL
O3460-11	A46M6	29485	5.3211 -		SFAM01.1	PPB	7/12/2023 18:18	SMPL
O3460-12	A46M7	1330	0.1673 -		SFAM01.1	PPB	7/12/2023 18:20	SMPL
O3460-13	A46N0	1822	0.2573 -		SFAM01.1	PPB	7/12/2023 18:23	SMPL
O3460-14	A46M4	1750	0.2441 -		SFAM01.1	PPB	7/12/2023 18:25	SMPL
O3460-15	A46M8	11271	1.987 -		SFAM01.1	PPB	7/12/2023 18:27	SMPL
O3460-16	A46M9	12577	2.2261 -		SFAM01.1	PPB	7/12/2023 18:29	SMPL
O3460-17	A46N1	18885	3.3808 -		SFAM01.1	PPB	7/12/2023 18:32	SMPL
O3460-18	A46N2	24264	4.3654 -		SFAM01.1	PPB	7/12/2023 18:34	SMPL
O3460-19	A46K5	27591	4.9744 -		SFAM01.1	PPB	7/12/2023 18:36	SMPL
O3460-20	A46K5D	27811	5.0147 -		SFAM01.1	PPB	7/12/2023 18:41	SMPL
O3460-21	A46K5S	42888	7.7746 -		SFAM01.1	PPB	7/12/2023 18:43	SMPL
CCV026	CCV026	28359	5.115 -		SFAM01.1	PPB	7/12/2023 18:46	SMPL
CCB026	CCB026	182	-0.0429 -		SFAM01.1	PPB	7/12/2023 18:48	SMPL
O3460-22	A46L4	2538	0.3884 -		SFAM01.1	PPB	7/12/2023 18:50	SMPL
O3460-23	A46L5	1329	0.1671 -		SFAM01.1	PPB	7/12/2023 18:52	SMPL
O3460-24	A46L6	361	-0.0101 -		SFAM01.1	PPB	7/12/2023 18:55	SMPL
CCV027	CCV027	28306	5.1053 -		SFAM01.1	PPB	7/12/2023 18:57	SMPL
CCB027	CCB027	156	-0.0476 -		SFAM01.1	PPB	7/12/2023 18:59	SMPL