



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126415

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/13/2023	08:20
2	CAL2	1	Water	NA	NA	20.8	7.00	07/13/2023	08:21
3	CAL3	1	Water	NA	NA	20.8	10.02	07/13/2023	08:22
4	ICV	1	Water	NA	NA	20.8	6.99	07/13/2023	08:25
5	CCV1	1	Water	NA	NA	20.7	2.01	07/13/2023	08:30
6	O3582-01	1	Water	NA	NA	25.6	4.69	07/13/2023	08:35
7	O3582-02	1	Water	NA	NA	26.0	4.82	07/13/2023	08:40
8	O3585-01	1	Water	NA	NA	21.1	6.22	07/13/2023	08:45
9	O3585-01DUP	1	Water	NA	NA	21.1	6.23	07/13/2023	08:46
10	CCV2	1	Water	NA	NA	20.6	12.02	07/13/2023	08:50

WORKLIST(Hardcopy Internal Chain)

VB 126415

WorkList Name : PH O3582 W

WorkList ID : 171830

Department : Wet-Chemistry

Date : 07-13-2023 08:04:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3582-01	MAIN-CELL	Water	pH	Cool 4 deg C	STER16	I11	07/11/2023	9040C
O3582-02	EX-CELL	Water	pH	Cool 4 deg C	STER16	I11	07/11/2023	9040C
O3585-01	50057	Water	pH	Cool 4 deg C	PSEG03	I11	07/12/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07-13-23 08:14

SD (CSC)

SD (CSC)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07-13-23

SD (CSC)

SD (CSC)