



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126416

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/13/2023	08:20
2	CAL2	1	Water	NA	NA	20.8	7.00	07/13/2023	08:21
3	CAL3	1	Water	NA	NA	20.8	10.02	07/13/2023	09:22
4	ICV	1	Water	NA	NA	20.8	6.99	07/13/2023	08:25
5	CCV1	1	Water	NA	NA	20.6	2.02	07/13/2023	09:00
6	O3586-01	1	Solid	20.02	20	21.7	7.21	07/13/2023	09:05
7	O3586-01DUP	1	Solid	20.03	20	21.8	7.22	07/13/2023	09:06
8	O3586-02	1	Solid	20.05	20	22.3	7.81	07/13/2023	09:10
9	O3586-03	1	Solid	20.02	20	22.3	6.87	07/13/2023	09:14
10	CCV2	1	Water	NA	NA	20.7	12.02	07/13/2023	09:15

WORKLIST(Hardcopy Internal Chain)

126416

WorkList Name : PH O3586 S WorkList ID : 171829 Department : Wet-Chemistry Date : 07-13-2023 08:04:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3586-01	C D1-3	Solid	pH	Cool 4 deg C	PSEG03	I11	07/12/2023	9045D
O3586-02	C D4-6	Solid	pH	Cool 4 deg C	PSEG03	I11	07/12/2023	9045D
O3586-03	C D7-9	Solid	pH	Cool 4 deg C	PSEG03	I11	07/12/2023	9045D

Date/Time 07-13-23 08:10

Raw Sample Received by: JDC (w/c)

Raw Sample Relinquished by: JDC (w/c)

Date/Time 07-13-23 11:00

Raw Sample Received by: JDC (w/c)

Raw Sample Relinquished by: JDC (w/c)