



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 7/14/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 14:15

In Date: 07/13/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:47

Out Date: 07/14/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB126422

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3603-01	MH0FD5	1	1.15	8.79	9.94	8.77	86.7	
O3603-02	MH0FD6	2	1.12	8.75	9.87	8.11	79.9	
O3603-03	MH0FD7	3	1.13	8.68	9.81	7.99	79.0	
O3603-04	MH0FD8	4	1.16	8.66	9.82	7.66	75.1	
O3603-05	MH0FD9	5	1.15	8.42	9.57	7.79	78.9	
O3603-06	MH0FE0	6	1.17	8.57	9.74	6.3	59.9	
O3603-07	MH0FE1	7	1.16	8.78	9.94	7.35	70.5	
O3603-08	MH0FE2	8	1.16	8.50	9.66	7.25	71.6	
O3603-09	MH0FE2D	9	1.16	8.50	9.66	7.25	71.6	
O3603-10	MH0FE2S	10	1.16	8.50	9.66	7.25	71.6	
O3603-11	MH0FE3	11	1.19	8.53	9.72	7.52	74.2	
O3603-12	MH0FE4	12	1.15	8.82	9.97	8.18	79.7	
O3603-13	MH0FF5	13	1.12	8.39	9.51	8.7	90.3	
O3603-14	MH0FF6	14	1.19	8.46	9.65	8.57	87.2	
O3603-15	MH0FF7	15	1.19	8.80	9.99	9.4	93.3	
O3603-16	MH0FF8	16	1.18	8.38	9.56	6.75	66.5	
O3603-17	MH0FF9	17	1.15	8.82	9.97	8.64	84.9	
O3603-18	MH0FG0	18	1.18	8.76	9.94	6.82	64.4	
O3603-19	MH0FG1	19	1.19	8.61	9.8	8.2	81.4	
O3603-20	MH0FG2	20	1.18	8.33	9.51	6.74	66.7	
O3603-21	MH0FG3	21	1.15	8.83	9.98	8.73	85.8	
O3603-22	MH0FG4	22	1.12	8.83	9.95	8.09	78.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$