


LB126424

SFAM01.1

INSTRUMENT ID: (V)



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.040	-0.040	220	0.000	220					
0.05	0.050					302	0				
0.20	0.200	0.172	-0.028	1377	0.0 %	1377					
2.50	2.500	2.518	0.018	14165	0.0 %	14165					
5.00	5.000	5.029	0.029	27856	0.0 %	27856					
7.50	7.500	7.680	0.180	42306	0.0 %	42306					
10.0	10.000	9.840	-0.160	54084	0.0 %	54084					

%D
= 14
1
1
2
-2

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	220	-		0 SFAM01.1	PPB	7/13/2023 10:52	Std
	0.2 S01	1377	-		0.2 SFAM01.1	PPB	7/13/2023 10:54	Std
	2.5 S02	14165	-		2.5 SFAM01.1	PPB	7/13/2023 10:56	Std
	5 S03	27856	-		5 SFAM01.1	PPB	7/13/2023 10:59	Std
	7.5 S04	42306	-		7.5 SFAM01.1	PPB	7/13/2023 11:01	Std
	10 S05	54084	-		10 SFAM01.1	PPB	7/13/2023 11:03	Std
ICV068	ICV068	22225	3.9965 -		SFAM01.1	PPB	7/13/2023 11:12	SMPL
ICB068	ICB068	179	-0.0474 -		SFAM01.1	PPB	7/13/2023 11:14	SMPL
CCV028	CCV028	28121	5.078 -		SFAM01.1	PPB	7/13/2023 11:16	SMPL
CCB028	CCB028	175	-0.0482 -		SFAM01.1	PPB	7/13/2023 11:18	SMPL
PB154175BL	PBS175	304	-0.0245 -		SFAM01.1	PPB	7/13/2023 11:20	SMPL
O3480-03	A46N3	1265	0.1518 -		SFAM01.1	PPB	7/13/2023 11:23	SMPL
O3480-04	A46N4	11423	2.0151 -		SFAM01.1	PPB	7/13/2023 11:25	SMPL
O3480-05	A46N5	15810	2.8198 -		SFAM01.1	PPB	7/13/2023 11:27	SMPL
O3480-06	A46N6	1129	0.1268 -		SFAM01.1	PPB	7/13/2023 11:30	SMPL
O3480-07	A46N8	12895	2.2851 -		SFAM01.1	PPB	7/13/2023 11:32	SMPL
O3480-08	A46N9	38350	6.9542 -		SFAM01.1	PPB	7/13/2023 11:34	SMPL
O3480-09	A46P0	632	0.0357 -		SFAM01.1	PPB	7/13/2023 11:36	SMPL
O3480-10	A46P3	3806	0.6179 -		SFAM01.1	PPB	7/13/2023 11:39	SMPL
O3480-11	A46P4	4179	0.6863 -		SFAM01.1	PPB	7/13/2023 11:41	SMPL
O3480-12	A46P7	9852	1.7269 -		SFAM01.1	PPB	7/13/2023 11:43	SMPL
O3480-13	A46P8	28829	5.2078 -		SFAM01.1	PPB	7/13/2023 11:45	SMPL
O3480-14	A46P9	1142	0.1292 -		SFAM01.1	PPB	7/13/2023 11:48	SMPL
O3480-15	A46Q6	25503	4.5977 -		SFAM01.1	PPB	7/13/2023 11:50	SMPL
O3480-16	A46Q7	35263	6.388 -		SFAM01.1	PPB	7/13/2023 11:52	SMPL
O3480-17	A46Q8	590	0.028 -		SFAM01.1	PPB	7/13/2023 11:54	SMPL
O3480-18	A46Q0	24507	4.415 -		SFAM01.1	PPB	7/13/2023 11:57	SMPL
O3480-19	A46Q0D	27261	4.9202 -		SFAM01.1	PPB	7/13/2023 11:59	SMPL
O3480-20	A46Q0S	40226	7.2984 -		SFAM01.1	PPB	7/13/2023 12:01	SMPL
O3480-21	A46S5	2141	0.3125 -		SFAM01.1	PPB	7/13/2023 12:03	SMPL
CCV029	CCV029	28204	5.0932 -		SFAM01.1	PPB	7/13/2023 12:06	SMPL
CCB029	CCB029	156	-0.0516 -		SFAM01.1	PPB	7/13/2023 12:08	SMPL
O3480-22	A46S6	2121	0.3088 -		SFAM01.1	PPB	7/13/2023 12:10	SMPL
O3480-23	A46S8	10792	1.8993 -		SFAM01.1	PPB	7/13/2023 12:13	SMPL
O3480-24	A46S9	12802	2.268 -		SFAM01.1	PPB	7/13/2023 12:15	SMPL
PB154176BL	PBS176	368	-0.0128 -		SFAM01.1	PPB	7/13/2023 12:17	SMPL
CCV030	CCV030	28096	5.0734 -		SFAM01.1	PPB	7/13/2023 12:54	SMPL
CCB030	CCB030	160	-0.0509 -		SFAM01.1	PPB	7/13/2023 12:57	SMPL
O3489-01	A46N7	700	0.0481 -		SFAM01.1	PPB	7/13/2023 12:59	SMPL
O3489-02	A46P1	1079	0.1177 -		SFAM01.1	PPB	7/13/2023 13:01	SMPL
O3489-03	A46P2	658	0.0404 -		SFAM01.1	PPB	7/13/2023 13:03	SMPL
O3489-04	A46P5	427	-0.0019 -		SFAM01.1	PPB	7/13/2023 13:06	SMPL
O3489-05	A46R2	3733	0.6045 -		SFAM01.1	PPB	7/13/2023 13:08	SMPL
O3489-06	A46R3	35513	6.4339 -		SFAM01.1	PPB	7/13/2023 13:10	SMPL
O3489-07	A46R4	31754	5.7443 -		SFAM01.1	PPB	7/13/2023 13:13	SMPL
O3489-08	A46R5	617	0.0329 -		SFAM01.1	PPB	7/13/2023 13:15	SMPL

INSTRUMENT ID : CV1

O3489-09	A46R6	368	-0.0128 -	SFAM01.1	PPB	7/13/2023 13:17 SMPL
O3489-10	A46Q3	25931	4.6762 -	SFAM01.1	PPB	7/13/2023 13:19 SMPL
O3489-11	A46Q4	602	0.0302 -	SFAM01.1	PPB	7/13/2023 13:22 SMPL
O3489-12	A46Q5	408	-0.0054 -	SFAM01.1	PPB	7/13/2023 13:24 SMPL
O3489-13	A46Q9	15688	2.7974 -	SFAM01.1	PPB	7/13/2023 13:26 SMPL
O3489-14	A46R0	32030	5.795 -	SFAM01.1	PPB	7/13/2023 13:28 SMPL
O3489-15	A46R1	436	-0.0003 -	SFAM01.1	PPB	7/13/2023 13:31 SMPL
CCV031	CCV031	28028	5.0609 -	SFAM01.1	PPB	7/13/2023 13:33 SMPL
CCB031	CCB031	153	-0.0522 -	SFAM01.1	PPB	7/13/2023 13:35 SMPL
O3489-16	A46Q1	24926	4.4919 -	SFAM01.1	PPB	7/13/2023 13:37 SMPL
O3489-17	A46Q1D	25241	4.5497 -	SFAM01.1	PPB	7/13/2023 13:40 SMPL
O3489-18	A46Q1S	38703	7.019 -	SFAM01.1	PPB	7/13/2023 13:42 SMPL
O3489-21	A46R8	3358	0.5357 -	SFAM01.1	PPB	7/13/2023 13:44 SMPL
O3489-22	A46S4	8675	1.511 -	SFAM01.1	PPB	7/13/2023 13:46 SMPL
CCV032	CCV032	28103	5.0747 -	SFAM01.1	PPB	7/13/2023 13:49 SMPL
CCB032	CCB032	151	-0.0526 -	SFAM01.1	PPB	7/13/2023 13:51 SMPL