



Analytical Summary Report

Analysis Method: SM5220 D
Parameter: COD
Run Number: LB126430

ANALYST: Iwona
SUPERVISOR REVIEW BY: Sohil

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP102919
COD calibration std. 100 ppm	WP102916
COD calibration std. 10 ppm	WP102914
COD calibration std. 150 ppm	WP102917
COD calibration std. 50 ppm	WP102915
COD calibration std. 0 ppm	WP102913
COD Digestion Vials Low Level 0-150Mg/L	W3024
COD ICV-LCS std, 50ppm	WP103189
COD CCV std, 50ppm	WP103188
COD LOD std, 5ppm	WP103235
COD LOQ std, 10.0PPM	WP103236

Temp In (C): 147	Date In: 07/12/2023	Time In: 14:05
Temp Out (C): 151	Date Out: 07/12/2023	Time Out: 16:05

Intercept: 0.4547 Slope: 0.9991 Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.455		06/14/2023	14:15
2	CAL2	10	1	Water	11.000	10.555	5.5	06/14/2023	14:15
3	CAL3	50	1	Water	51.000	50.591	1.2	06/14/2023	14:16
4	CAL4	100	1	Water	99.000	98.634	-1.4	06/14/2023	14:16
5	CAL5	150	1	Water	151.000	150.681	0.5	06/14/2023	14:17



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126430

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	49.000	48.589	06/14/2023	14:17
2	ICB		NA	NA	1	Water	1.000	0.546	06/14/2023	14:18
3	CCV1	50	NA	NA	1	Water	49.000	48.589	07/13/2023	10:50
4	CCB1		NA	NA	1	Water	1.000	0.546	07/13/2023	10:50
5	LB126430BL		1	50	1	Solid	0.000	-0.455	07/13/2023	10:51
6	LB126430BS	50	1	50	1	Solid	49.000	48.589	07/13/2023	10:51
7	O3572-01		1	50	1	Solid	6.000	5.550	07/13/2023	10:52
8	O3572-02		1	50	1	Solid	10.000	9.554	07/13/2023	10:52
9	CCV2	50	NA	NA	1	Water	50.000	49.590	07/13/2023	10:53
10	CCB2		NA	NA	1	Water	1.000	0.546	07/13/2023	10:53