

LB126438


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5732E



A= 0.0000e+000

B= 1.7592e-004 slope

C= -3.2382e-002 y-intercept

Rho= 0.9999469

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/H
0.00	0.000	-0.011	-0.011	122	0.000	122					
0.05	0.050					302	0				
0.20	0.200	0.248	0.048	1596	0.0 %	1596					
2.50	2.500	2.491	-0.009	14344	0.0 %	14344					
5.00	5.000	4.966	-0.034	28411	0.0 %	28411					
7.50	7.500	7.453	-0.047	42552	0.0 %	42552					
10.0	10.000	10.053	0.053	57328	0.0 %	57328					

G/H
 11
 24
 11
 11
 1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	122	-		0 SFAM01.1	PPB	7/13/2023 16:23	Std
	0.2 S01	1596	-		0.2 SFAM01.1	PPB	7/13/2023 16:25	Std
	2.5 S02	14344	-		2.5 SFAM01.1	PPB	7/13/2023 16:27	Std
	5 S03	28411	-		5 SFAM01.1	PPB	7/13/2023 16:31	Std
	7.5 S04	42552	-		7.5 SFAM01.1	PPB	7/13/2023 16:34	Std
	10 S05	57328	-		10 SFAM01.1	PPB	7/13/2023 16:36	Std
ICV070	ICV070	22248	3.8814 -		SFAM01.1	PPB	7/13/2023 16:39	SMPL
ICB070	ICB070	90	-0.0165 -		SFAM01.1	PPB	7/13/2023 16:41	SMPL
CCV036	CCV036	28342	4.9535 -		SFAM01.1	PPB	7/13/2023 16:43	SMPL
CCB036	CCB036	96	-0.0155 -		SFAM01.1	PPB	7/13/2023 16:45	SMPL
PB154186BL	PBS186	115	-0.0122 -		SFAM01.1	PPB	7/13/2023 16:48	SMPL
O3602-01	MH0FC5	2453	0.3991 -		SFAM01.1	PPB	7/13/2023 16:50	SMPL
O3602-02	MH0FC6	935	0.1321 -		SFAM01.1	PPB	7/13/2023 16:52	SMPL
O3602-03	MH0FC7	1209	0.1803 -		SFAM01.1	PPB	7/13/2023 16:54	SMPL
O3602-04	MH0FC7D	1280	0.1928 -		SFAM01.1	PPB	7/13/2023 16:57	SMPL
O3602-05	MH0FC7S	14552	2.5276 -		SFAM01.1	PPB	7/13/2023 16:59	SMPL
O3602-06	MH0FC8	961	0.1367 -		SFAM01.1	PPB	7/13/2023 17:01	SMPL
O3602-07	MH0FC9	1214	0.1812 -		SFAM01.1	PPB	7/13/2023 17:04	SMPL
O3602-08	MH0FD0	10393	1.7959 -		SFAM01.1	PPB	7/13/2023 17:06	SMPL
O3602-09	MH0FD1	2499	0.4072 -		SFAM01.1	PPB	7/13/2023 17:08	SMPL
O3602-10	MH0FD2	576	0.0689 -		SFAM01.1	PPB	7/13/2023 17:10	SMPL
O3602-11	MH0FD3	441	0.0452 -		SFAM01.1	PPB	7/13/2023 17:13	SMPL
O3602-12	MH0FD4	1673	0.2619 -		SFAM01.1	PPB	7/13/2023 17:15	SMPL
O3602-13	MH0FE5	3036	0.5017 -		SFAM01.1	PPB	7/13/2023 17:17	SMPL
O3602-14	MH0FE6	2755	0.4523 -		SFAM01.1	PPB	7/13/2023 17:19	SMPL
O3602-15	MH0FE7	3808	0.6375 -		SFAM01.1	PPB	7/13/2023 17:22	SMPL
O3602-16	MH0FE8	4113	0.6912 -		SFAM01.1	PPB	7/13/2023 17:24	SMPL
O3602-17	MH0FE9	1583	0.2461 -		SFAM01.1	PPB	7/13/2023 17:26	SMPL
O3602-18	MH0FF0	904	0.1266 -		SFAM01.1	PPB	7/13/2023 17:28	SMPL
O3602-19	MH0FF1	728	0.0957 -		SFAM01.1	PPB	7/13/2023 17:31	SMPL
CCV037	CCV037	28429	4.9688 -		SFAM01.1	PPB	7/13/2023 17:33	SMPL
CCB037	CCB037	92	-0.0162 -		SFAM01.1	PPB	7/13/2023 17:35	SMPL
O3602-20	MH0FF2	2861	0.4709 -		SFAM01.1	PPB	7/13/2023 17:38	SMPL
O3602-21	MH0FF3	1566	0.2431 -		SFAM01.1	PPB	7/13/2023 17:40	SMPL
O3602-22	MH0FF4	1811	0.2862 -		SFAM01.1	PPB	7/13/2023 17:42	SMPL
PB154187BL	PBS187	620	0.0767 -		SFAM01.1	PPB	7/13/2023 17:44	SMPL
O3603-01	MH0FD5	1860	0.2948 -		SFAM01.1	PPB	7/13/2023 17:47	SMPL
O3603-02	MH0FD6	2027	0.3242 -		SFAM01.1	PPB	7/13/2023 17:49	SMPL
O3603-03	MH0FD7	2004	0.3202 -		SFAM01.1	PPB	7/13/2023 17:51	SMPL
O3603-04	MH0FD8	2611	0.4269 -		SFAM01.1	PPB	7/13/2023 17:53	SMPL
O3603-05	MH0FD9	2960	0.4883 -		SFAM01.1	PPB	7/13/2023 17:56	SMPL
O3603-06	MH0FE0	1785	0.2816 -		SFAM01.1	PPB	7/13/2023 17:58	SMPL
O3603-07	MH0FE1	1500	0.2315 -		SFAM01.1	PPB	7/13/2023 18:00	SMPL
O3603-08	MH0FE2	2542	0.4148 -		SFAM01.1	PPB	7/13/2023 18:02	SMPL
O3603-09	MH0FE2D	2524	0.4116 -		SFAM01.1	PPB	7/13/2023 18:05	SMPL
O3603-10	MH0FE2S	16272	2.8302 -		SFAM01.1	PPB	7/13/2023 18:07	SMPL

LB126438

INSTRUMENT ID : CV1

O3603-11	MH0FE3	3204	0.5313 -	SFAM01.1	PPB	7/13/2023 18:09 SMPL
O3603-12	MH0FE4	2755	0.4523 -	SFAM01.1	PPB	7/13/2023 18:11 SMPL
O3603-13	MH0FF5	28907	5.0529 -	SFAM01.1	PPB	7/13/2023 18:14 SMPL
O3603-14	MH0FF6	707	0.092 -	SFAM01.1	PPB	7/13/2023 18:16 SMPL
O3603-15	MH0FF7	1063	0.1546 -	SFAM01.1	PPB	7/13/2023 18:18 SMPL
O3603-16	MH0FF8	1759	0.2771 -	SFAM01.1	PPB	7/13/2023 18:20 SMPL
CCV038	CCV038	28446	4.9718 -	SFAM01.1	PPB	7/13/2023 18:23 SMPL
CCB038	CCB038	84	-0.0176 -	SFAM01.1	PPB	7/13/2023 18:25 SMPL
O3603-17	MH0FF9	1849	0.2929 -	SFAM01.1	PPB	7/13/2023 18:27 SMPL
O3603-18	MH0FG0	1281	0.193 -	SFAM01.1	PPB	7/13/2023 18:30 SMPL
O3603-19	MH0FG1	1514	0.234 -	SFAM01.1	PPB	7/13/2023 18:32 SMPL
O3603-20	MH0FG2	1159	0.1715 -	SFAM01.1	PPB	7/13/2023 18:34 SMPL
O3603-21	MH0FG3	975	0.1391 -	SFAM01.1	PPB	7/13/2023 18:36 SMPL
O3603-22	MH0FG4	438	0.0447 -	SFAM01.1	PPB	7/13/2023 18:39 SMPL
CCV039	CCV039	28262	4.9394 -	SFAM01.1	PPB	7/13/2023 18:41 SMPL
CCB039	CCB039	103	-0.0143 -	SFAM01.1	PPB	7/13/2023 18:43 SMPL