



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB126444

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/14/2023	09:25
2	CAL2	1	Water	NA	NA	20.7	7.00	07/14/2023	09:26
3	CAL3	1	Water	NA	NA	20.7	10.03	07/14/2023	09:30
4	ICV	1	Water	NA	NA	20.7	7.02	07/14/2023	09:33
5	CCV1	1	Water	NA	NA	20.7	2.01	07/14/2023	09:34
6	O3615-08	1	Water	NA	NA	20.6	7.45	07/14/2023	09:40
7	O3616-01	1	Water	NA	NA	25.2	9.94	07/14/2023	09:50
8	O3620-01	1	Water	NA	NA	21.9	7.51	07/14/2023	10:00
9	O3620-02	1	Water	NA	NA	21.7	8.94	07/14/2023	10:05
10	O3620-03	1	Water	NA	NA	21.8	8.42	07/14/2023	10:10
11	O3620-04	1	Water	NA	NA	21.9	8.38	07/14/2023	10:11
12	O3620-04DUP	1	Water	NA	NA	21.8	8.39	07/14/2023	10:12
13	CCV2	1	Water	NA	NA	20.7	12.02	07/14/2023	10:20

UB 126444

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O3620 WorkList ID : 171881 Department : Wet-Chemistry Date : 07-14-2023 09:08:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3615-08	A 2880	Water	pH	Cool 4 deg C	PSEG03	I11	07/13/2023	9040C
O3616-01	FRAC-TOTE-COMP	Water	pH	Cool 4 deg C	PSEG03	I11	07/13/2023	9040C
O3620-01	C 713-A	Water	pH	Cool 4 deg C	PSEG03	I11	07/13/2023	9040C
O3620-02	C 713-B	Water	pH	Cool 4 deg C	PSEG03	I11	07/13/2023	9040C
O3620-03	C 713-C	Water	pH	Cool 4 deg C	PSEG03	I11	07/13/2023	9040C
O3620-04	C 713-D	Water	pH	Cool 4 deg C	PSEG03	I11	07/13/2023	9040C

Date/Time 07-14-23 09:15
Raw Sample Received by: SO/DOC
Raw Sample Relinquished by: CP SM

Date/Time 07-14-23 11:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: SO/DOC