



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126453

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	07/14/2023	12:30
2	O3584-01		1	12	100.00	>212.0	>212.0	07/14/2023	13:00
3	O3615-01		1	11	100.00	>212.0	>212.0	07/14/2023	13:30
4	O3619-04		1	3	8.00	46.4	46.4	07/14/2023	14:00
5	O3619-04DUP		1	3	8.00	46.4	46.4	07/14/2023	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

25126453

WorkList Name : fp s-7-13

WorkList ID : 171824

Department : Wet-Chemistry

Date : 07-13-2023 07:58:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3584-01 A	0605	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	07/12/2023	1010B
O3615-01 A	2881	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	07/13/2023	1010B
O3619-04 A	D3135	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	07/13/2023	1010B

Date/Time 07/14/2023 12:20
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RA CWG

Date/Time 07/14/2023 14:30
 Raw Sample Received by: JG CWG
 Raw Sample Relinquished by: RA CWG