



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126459

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP102919
COD calibration std. 100 ppm	WP102916
COD calibration std. 10 ppm	WP102914
COD calibration std. 150 ppm	WP102917
COD calibration std. 50 ppm	WP102915
COD calibration std. 0 ppm	WP102913
COD Digestion Vials Low Level 0-150Mg/L	W3024
COD ICV-LCS std, 50ppm	WP103189
COD CCV std, 50ppm	WP103188

Temp In (C): 148	Date In: 07/14/2023	Time In: 11:00
Temp Out (C): 151	Date Out: 07/14/2023	Time Out: 13:00

Intercept: 0.4547

Slope: 0.9991

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.455		06/14/2023	14:15
2	CAL2	10	1	Water	11.000	10.555	5.5	06/14/2023	14:15
3	CAL3	50	1	Water	51.000	50.591	1.2	06/14/2023	14:16
4	CAL4	100	1	Water	99.000	98.634	-1.4	06/14/2023	14:16
5	CAL5	150	1	Water	151.000	150.681	0.5	06/14/2023	14:17



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126459

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	49.000	48.589	06/14/2023	14:17
2	ICB		NA	NA	1	Water	1.000	0.546	06/14/2023	14:18
3	CCV1	50	NA	NA	1	Water	51.000	50.591	07/13/2023	15:10
4	CCB1		NA	NA	1	Water	0.000	-0.455	07/13/2023	15:10
5	LB126459BL		NA	NA	1	Water	1.000	0.546	07/13/2023	15:11
6	LB126459BS	50	NA	NA	1	Water	50.000	49.590	07/13/2023	15:11
7	O3546-01		NA	NA	1	Water	4.000	3.548	07/13/2023	15:12
8	O3546-01DUP		NA	NA	1	Water	4.000	3.548	07/13/2023	15:12
9	O3546-01MS	50	NA	NA	1	Water	50.000	49.590	07/13/2023	15:13
10	O3546-01MSD	50	NA	NA	1	Water	51.000	50.591	07/13/2023	15:13
11	O3546-03		NA	NA	1	Water	2.000	1.547	07/13/2023	15:14
12	CCV2	50	NA	NA	1	Water	49.000	48.589	07/13/2023	15:14
13	CCB2		NA	NA	1	Water	1.000	0.546	07/13/2023	15:15

WORKLIST(Hardcopy Internal Chain)

LB 126459

WorkList Name : COD-071323

WorkList ID : 171900

Department : Wet-Chemistry

Date : 07-13-2023 10:10:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3546-01	1675	Water	COD	Conc H2SO4 to pH < 2	PSEG03	I11	07/06/2023	SM5220 D
O3546-03	1676	Water	COD	Conc H2SO4 to pH < 2	PSEG03	I11	07/06/2023	SM5220 D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/13/23 10:30
12/51
78(60)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/13/23 14:20
78(60)
12/54