



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126460

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/14/2023	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	07/14/2023	16:26
3	CAL3	1	Water	NA	NA	20.7	10.02	07/14/2023	16:27
4	ICV	1	Water	NA	NA	20.7	7.02	07/14/2023	16:30
5	CCV1	1	Water	NA	NA	20.7	2.02	07/14/2023	16:34
6	O3637-01	1	Water	NA	NA	22.2	6.87	07/14/2023	16:37
7	O3637-01DUP	1	Water	NA	NA	22.3	6.88	07/14/2023	16:38
8	CCV2	1	Water	NA	NA	20.7	12.02	07/14/2023	16:40

VB 126460

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O3637

WorkList ID : 171904

Department : Wet-Chemistry

Date : 07-14-2023 16:13:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3637-01 L	A508	Water	pH	Cool 4 deg C	PSEG03	I11	07/14/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07-14-23 16:20

SDC/

SDC/

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07-14-23

SDC/

SDC/