

LB126474


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46950



A= 0.0000e+000

B= 2.1168e-004

C= -5.0945e-002

Rho= 0.9995798

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.033	0.033	398	0.000	398	0				7
0.2	0.200	0.215	0.015	1255	0.0 %	1255					-3
2.5	2.500	2.418	-0.082	11661	0.0 %	11661					-1
5.0	5.000	4.937	-0.063	23564	0.0 %	23564					3
7.5	7.500	7.710	0.210	36661	0.0 %	36661					-1
10.0	10.000	9.888	-0.112	46950	0.0 %	46950					

LB126474 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jul 2023 10:34:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	398	-	0.0000	7470A	PPB	17 Jul 2023 10:56:15	S	Std
0.2 - 1	502	1255	-	0.2000	7470A	PPB	17 Jul 2023 10:58:32	S	Std
2.5 - 1	52.5	11661	-	2.5000	7470A	PPB	17 Jul 2023 11:00:49	S	Std
5.0 - 1	55	23564	-	5.0000	7470A	PPB	17 Jul 2023 11:03:06	S	Std
7.5 - 1	57.5	36661	-	7.5000	7470A	PPB	17 Jul 2023 11:08:01	S	Std
10.0 - 1	510	46950	-	10.0000	7470A	PPB	17 Jul 2023 11:10:16	S	Std
ICV72 - 1	ICV72	17899	3.7380	-	7470A	PPB	17 Jul 2023 11:15:40	U	SMPL
ICB72 - 1	ICB72	119	-0.0258	-	7470A	PPB	17 Jul 2023 11:17:54	U	SMPL
CCV45 - 1	CCV45	23973	5.0238	-	7470A	PPB	17 Jul 2023 11:20:09	U	SMPL
CCB45 - 1	CCB45	116	-0.0264	-	7470A	PPB	17 Jul 2023 11:22:24	U	SMPL
CRA - 1	CRA	1258	0.2154	-	7470A	PPB	17 Jul 2023 11:24:39	U	SMPL
HighStd - 1	HighStd	47263	9.9539	-	7470A	PPB	17 Jul 2023 11:26:54	U	SMPL
ChkStd - 1	ChkStd	33480	7.0363	-	7470A	PPB	17 Jul 2023 11:29:09	U	SMPL
PB154201BL - 1	PBW	-5	-0.0520	-	7470A	PPB	17 Jul 2023 11:31:24	U	SMPL
PB154201BS - 1	LCSW	21644	4.5308	-	7470A	PPB	17 Jul 2023 11:35:51	U	SMPL
O3580-01 - 1	WASTE-1	14	-0.0480	-	7470A	PPB	17 Jul 2023 11:38:07	U	SMPL
O3580-01DUP - 1	WASTE-1DUP	58	-0.0387	-	7470A	PPB	17 Jul 2023 11:40:23	U	SMPL
O3580-01MS - 1	WASTE-1MS	20495	4.2875	-	7470A	PPB	17 Jul 2023 11:47:49	U	SMPL
O3580-01MSD - 1	WASTE-1MSD	20284	4.2429	-	7470A	PPB	17 Jul 2023 11:52:42	U	SMPL
O3599-01 - 1	IDW-7-11-23	50	-0.0404	-	7470A	PPB	17 Jul 2023 11:54:57	U	SMPL
CCV46 - 1	CCV46	24046	5.0392	-	7470A	PPB	17 Jul 2023 11:57:13	U	SMPL
CCB46 - 1	CCB46	113	-0.0270	-	7470A	PPB	17 Jul 2023 11:59:27	U	SMPL
O3615-01 - 1	2881	1515	0.2698	-	7470A	PPB	17 Jul 2023 12:01:42	U	SMPL
O3615-02 - 1	2884	84	-0.0332	-	7470A	PPB	17 Jul 2023 12:03:57	U	SMPL
O3615-03 - 1	2885-2887	150	-0.0192	-	7470A	PPB	17 Jul 2023 12:06:12	U	SMPL
O3619-03 - 1	D3121	35	-0.0435	-	7470A	PPB	17 Jul 2023 12:08:28	U	SMPL
O3619-04 - 1	D3135	4733	0.9510	-	7470A	PPB	17 Jul 2023 12:10:44	U	SMPL
PB154202BL - 1	PBW	-4	-0.0518	-	7470A	PPB	17 Jul 2023 12:15:12	U	SMPL
PB154202BS - 1	LCSW	20795	4.3510	-	7470A	PPB	17 Jul 2023 12:20:13	U	SMPL
O3599-02 - 1	FIELD-BLANK	-140	-0.0806	-	7470A	PPB	17 Jul 2023 12:22:30	U	SMPL
O3616-01 - 1	FRAC-TOTE-COMP	3659	0.7236	-	7470A	PPB	17 Jul 2023 12:29:18	U	SMPL
O3616-01DUP - 1	FRAC-TOTE-COMP DUP	3692	0.7306	-	7470A	PPB	17 Jul 2023 12:31:33	U	SMPL
CCV47 - 1	CCV47	23582	4.9410	-	7470A	PPB	17 Jul 2023 12:33:48	U	SMPL
CCB47 - 1	CCB47	66	-0.0370	-	7470A	PPB	17 Jul 2023 12:36:04	U	SMPL
O3616-01MS - 1	FRAC-TOTE-COMPMS	22198	4.6480	-	7470A	PPB	17 Jul 2023 12:41:18	U	SMPL
O3616-01MSD - 1	FRAC-TOTE-COMPMSD	19887	4.1588	-	7470A	PPB	17 Jul 2023 12:43:33	U	SMPL
PB154165TB - 1	PB154165TB	-29	-0.0571	-	7470A	PPB	17 Jul 2023 12:45:49	U	SMPL
O3580-01LX5 - 1		39	-0.0427	-	7470A	PPB	17 Jul 2023 12:48:04	U	SMPL
O3580-01A - 1		18082	3.7767	-	7470A	PPB	17 Jul 2023 12:50:20	U	SMPL
O3616-01LX5 - 1		1025	0.1660	-	7470A	PPB	17 Jul 2023 12:52:36	U	SMPL
O3616-01A - 1		27281	5.7240	-	7470A	PPB	17 Jul 2023 12:54:52	U	SMPL
CCV48 - 1	CCV48	22904	4.7975	-	7470A	PPB	17 Jul 2023 12:57:09	U	SMPL
CCB48 - 1	CCB48	263	0.0047	-	7470A	PPB	17 Jul 2023 12:59:27	U	SMPL