


LB126479

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.5469e-004

C= 7.3743e-002

Rho= 0.9998520

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.093	0.093	126	0.000	126					0
0.2	0.200	0.203	0.003	834	0.0 %	834					2
2.5	2.500	2.421	-0.079	15172	0.0 %	15172					3
5.0	5.000	4.917	-0.083	31307	0.0 %	31307					2
7.5	7.500	7.531	0.031	48206	0.0 %	48206					0
10.0	10.000	10.036	0.036	64402	0.0 %	64402					0

LB126479 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jul 2023 14:32:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	126	-	0.0000	7471B	PPB	17 Jul 2023 14:37:23	S	Std
0.2 - 1	50.2	834	-	0.2000	7471B	PPB	17 Jul 2023 14:39:40	S	Std
2.5 - 1	52.5	15172	-	2.5000	7471B	PPB	17 Jul 2023 14:41:57	S	Std
5.0 - 1	55	31307	-	5.0000	7471B	PPB	17 Jul 2023 14:44:13	S	Std
7.5 - 1	57.5	48206	-	7.5000	7471B	PPB	17 Jul 2023 14:46:30	S	Std
10.0 - 1	510	64402	-	10.0000	7471B	PPB	17 Jul 2023 14:56:34	S	Std
ICV73 - 1	ICV73	24666	3.8893	-	7471B	PPB	17 Jul 2023 14:59:32	U	SMPL
ICB73 - 1	ICB73	33	0.0788	-	7471B	PPB	17 Jul 2023 15:01:46	U	SMPL
CCV49 - 1	CCV49	31601	4.9621	-	7471B	PPB	17 Jul 2023 15:04:01	U	SMPL
CCB49 - 1	CCB49	56	0.0824	-	7471B	PPB	17 Jul 2023 15:06:16	U	SMPL
CRA - 1	CRA	815	0.1998	-	7471B	PPB	17 Jul 2023 15:08:31	U	SMPL
HighStd - 1	HighStd	65119	10.1469	-	7471B	PPB	17 Jul 2023 15:10:46	U	SMPL
ChkStd - 1	ChkStd	43215	6.7586	-	7471B	PPB	17 Jul 2023 15:13:01	U	SMPL
PB154225BL - 1	PBS	42	0.0802	-	7471B	PPB	17 Jul 2023 15:15:18	U	SMPL
PB154225BS - 1	LCSS	24749	3.9021	-	7471B	PPB	17 Jul 2023 15:17:35	U	SMPL
O3613-01 - 1	TR-4-071323	4965	0.8418	-	7471B	PPB	17 Jul 2023 15:19:51	U	SMPL
O3615-04 - 1	VNJ-210	7351	1.2109	-	7471B	PPB	17 Jul 2023 15:22:08	U	SMPL
O3615-04DUP - 1	VNJ-210DUP	6473	1.0750	-	7471B	PPB	17 Jul 2023 15:24:25	U	SMPL
O3615-04MS - 1	VNJ-210MS	33044	5.1853	-	7471B	PPB	17 Jul 2023 15:30:55	U	SMPL
O3615-04MSD - 1	VNJ-210MSD	29866	4.6937	-	7471B	PPB	17 Jul 2023 15:33:10	U	SMPL
CCV50 - 1	CCV50	31265	4.9101	-	7471B	PPB	17 Jul 2023 15:35:26	U	SMPL
CCB50 - 1	CCB50	53	0.0819	-	7471B	PPB	17 Jul 2023 15:37:40	U	SMPL
O3615-06 - 1	VNJ-204	9576	1.5550	-	7471B	PPB	17 Jul 2023 15:39:55	U	SMPL
O3617-01 - 1	VNJ-214	38471	6.0248	-	7471B	PPB	17 Jul 2023 15:42:10	U	SMPL
O3619-01 - 1	RT3471	3429	0.6042	-	7471B	PPB	17 Jul 2023 15:44:26	U	SMPL
O3621-01 - 1	OR-03-071323	3821	0.6648	-	7471B	PPB	17 Jul 2023 15:46:43	U	SMPL
O3625-01 - 1	RBR200018	10641	1.7198	-	7471B	PPB	17 Jul 2023 15:48:58	U	SMPL
O3632-01 - 1	SB-01-(2-4)	49267	7.6948	-	7471B	PPB	17 Jul 2023 15:51:14	U	SMPL
O3632-02 - 1	SB-02-(5-7)	1004	0.2291	-	7471B	PPB	17 Jul 2023 15:53:30	U	SMPL
O3632-03 - 1	SB-03-(7-9)	1601	0.3214	-	7471B	PPB	17 Jul 2023 15:55:48	U	SMPL
O3632-04 - 1	SB-04-(9-11)	1358	0.2838	-	7471B	PPB	17 Jul 2023 15:58:05	U	SMPL
O3632-06 - 1	SB-09-(3-5)	1375	0.2864	-	7471B	PPB	17 Jul 2023 16:00:20	U	SMPL
CCV51 - 1	CCV51	31078	4.8812	-	7471B	PPB	17 Jul 2023 16:02:35	U	SMPL
CCB51 - 1	CCB51	95	0.0884	-	7471B	PPB	17 Jul 2023 16:04:51	U	SMPL
O3632-07 - 1	SB-10-(9-11)	12616	2.0253	-	7471B	PPB	17 Jul 2023 16:07:06	U	SMPL
O3632-08 - 1	SB-12-(4-6)	1914	0.3698	-	7471B	PPB	17 Jul 2023 16:09:21	U	SMPL
PB154226BL - 1	PBS	90	0.0877	-	7471B	PPB	17 Jul 2023 16:14:16	U	SMPL
PB154226BS - 1	LCSS	25010	3.9425	-	7471B	PPB	17 Jul 2023 16:16:31	U	SMPL
O3597-01 - 1	RB-42-(1-3)	10173	1.6474	-	7471B	PPB	17 Jul 2023 16:18:47	U	SMPL
O3597-02 - 1	RB-42-(5-7)	4836	0.8218	-	7471B	PPB	17 Jul 2023 16:21:03	U	SMPL
O3597-03 - 1	RB-43-(0-2)	42485	6.6457	-	7471B	PPB	17 Jul 2023 16:23:20	U	SMPL
O3597-04 - 1	RB-43-(2-5)	4799	0.8161	-	7471B	PPB	17 Jul 2023 16:25:38	U	SMPL
O3597-05 - 1	RB-44-(0-2)	5835	0.9764	-	7471B	PPB	17 Jul 2023 16:27:55	U	SMPL
O3597-06 - 1	RB-44-(2-5)	21265	3.3632	-	7471B	PPB	17 Jul 2023 16:30:11	U	SMPL
CCV52 - 1	CCV52	31015	4.8714	-	7471B	PPB	17 Jul 2023 16:32:26	U	SMPL
CCB52 - 1	CCB52	15	0.0761	-	7471B	PPB	17 Jul 2023 16:34:41	U	SMPL
O3597-07 - 1	RB-45-(0-2)	12516	2.0098	-	7471B	PPB	17 Jul 2023 16:36:57	U	SMPL
O3597-08 - 1	RB-45-(2-5)	4639	0.7913	-	7471B	PPB	17 Jul 2023 16:39:12	U	SMPL
O3597-09 - 1	RB-46-(0-2)	25015	3.9433	-	7471B	PPB	17 Jul 2023 16:41:27	U	SMPL
O3597-10 - 1	RB-46-(2-5)	119653	18.5828	-	7471B	PPB	17 Jul 2023 16:43:43	U	SMPL
O3597-11 - 1	RB-47-(0-2)	21544	3.4064	-	7471B	PPB	17 Jul 2023 16:45:59	U	SMPL
O3597-12 - 1	RB-47-(2-5)	2117	0.4012	-	7471B	PPB	17 Jul 2023 16:48:18	U	SMPL
O3597-13 - 1	RB-48-(0-2)	37326	5.8477	-	7471B	PPB	17 Jul 2023 16:50:34	U	SMPL
O3597-14 - 1	RB-48-(2-5)	17715	2.8141	-	7471B	PPB	17 Jul 2023 16:52:51	U	SMPL
O3597-15 - 1	RB-49-(0-2)	32457	5.0945	-	7471B	PPB	17 Jul 2023 16:55:10	U	SMPL
O3597-16 - 1	RB-49-(2-5)	5446	0.9162	-	7471B	PPB	17 Jul 2023 16:57:26	U	SMPL
CCV53 - 1	CCV53	31403	4.9314	-	7471B	PPB	17 Jul 2023 16:59:41	U	SMPL
CCB53 - 1	CCB53	-20	0.0706	-	7471B	PPB	17 Jul 2023 17:01:57	U	SMPL
O3597-17 - 1	RB-50-(0-2)	55069	8.5923	-	7471B	PPB	17 Jul 2023 17:04:14	U	SMPL
O3597-18 - 1	RB-50-(2-5)	6878	1.1377	-	7471B	PPB	17 Jul 2023 17:06:29	U	SMPL
O3597-19 - 1	RB-51-(0-2)	136556	21.1975	-	7471B	PPB	17 Jul 2023 17:08:47	U	SMPL
O3597-20 - 1	RB-51-(2-5)	15488	2.4696	-	7471B	PPB	17 Jul 2023 17:11:03	U	SMPL
O3597-20DUP - 1	RB-51-(2-5)DUP	16367	2.6055	-	7471B	PPB	17 Jul 2023 17:13:21	U	SMPL
O3597-21 - 1	O3597-20MS	42110	6.5877	-	7471B	PPB	17 Jul 2023 17:15:37	U	SMPL
O3597-22 - 1	O3597-20MSD	42384	6.6301	-	7471B	PPB	17 Jul 2023 17:17:53	U	SMPL
O3615-04LX5 - 1		1132	0.2489	-	7471B	PPB	17 Jul 2023 17:20:16	U	SMPL
O3615-04A - 1		34482	5.4077	-	7471B	PPB	17 Jul 2023 17:22:34	U	SMPL
CCV54 - 1	CCV54	31399	4.9308	-	7471B	PPB	17 Jul 2023 17:24:49	U	SMPL
CCB54 - 1	CCB54	-13	0.0717	-	7471B	PPB	17 Jul 2023 17:27:08	U	SMPL
O3597-20LX5 - 1		3293	0.5831	-	7471B	PPB	17 Jul 2023 17:29:27	U	SMPL
O3597-20A - 1		37211	5.8299	-	7471B	PPB	17 Jul 2023 17:31:43	U	SMPL
O3597-10DLX5 - 1		25908	4.0814	-	7471B	PPB	17 Jul 2023 17:33:58	U	SMPL
O3597_19DLX5 - 1		29169	4.5859	-	7471B	PPB	17 Jul 2023 17:36:18	U	SMPL
CCV55 - 1	CCV55	31460	4.9403	-	7471B	PPB	17 Jul 2023 17:38:34	U	SMPL
CCB55 - 1	CCB55	-152	0.0502	-	7471B	PPB	17 Jul 2023 17:40:52	U	SMPL