

LB126492


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46777



A= 0.0000e+000

B= 2.1444e-004

C= 1.3149e-002

Rho= 0.9996713

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/σ)
0.0	0.000	0.037	0.037	109	0.000	109					1
0.2	0.200	0.193	-0.007	839	0.0 %	839					-4
2.5	2.500	2.412	-0.088	11185	0.0 %	11185					4
5.0	5.000	5.153	0.153	23968	0.0 %	23968					3
7.5	7.500	7.362	-0.138	34267	0.0 %	34267					2
10.0	10.000	10.044	0.044	46777	0.0 %	46777					0

LB126492 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jul 2023 11:43:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	109	-	0.0000	7470A	PPB	18 Jul 2023 11:59:16	S	Std
0.2 - 1	50.2	839	-	0.2000	7470A	PPB	18 Jul 2023 12:04:07	S	Std
2.5 - 1	52.5	11185	-	2.5000	7470A	PPB	18 Jul 2023 12:06:24	S	Std
5.0 - 1	55	23968	-	5.0000	7470A	PPB	18 Jul 2023 12:08:40	S	Std
7.5 - 1	57.5	34267	-	7.5000	7470A	PPB	18 Jul 2023 12:10:57	S	Std
10.0 - 1	510	46777	-	10.0000	7470A	PPB	18 Jul 2023 12:13:12	S	Std
ICV74 - 1	ICV74	19685	4.2345	-	7470A	PPB	18 Jul 2023 12:16:02	U	SMPL
ICB74 - 1	ICB74	-26	0.0076	-	7470A	PPB	18 Jul 2023 12:18:17	U	SMPL
CCV56 - 1	CCV56	22688	4.8785	-	7470A	PPB	18 Jul 2023 12:20:34	U	SMPL
CCB56 - 1	CCB56	-2	0.0127	-	7470A	PPB	18 Jul 2023 12:22:49	U	SMPL
CRA - 1	CRA	863	0.1982	-	7470A	PPB	18 Jul 2023 12:25:06	U	SMPL
HighStd - 1	HighStd	46543	9.9940	-	7470A	PPB	18 Jul 2023 12:27:20	U	SMPL
ChkStd - 1	ChkStd	31769	6.8258	-	7470A	PPB	18 Jul 2023 12:29:35	U	SMPL
PB154247BL - 1	PBW	4	0.0140	-	7470A	PPB	18 Jul 2023 12:31:53	U	SMPL
PB154247BS - 1	LCSW	16465	3.5440	-	7470A	PPB	18 Jul 2023 12:34:08	U	SMPL
O3625-02 - 1	RBR200018	73	0.0288	-	7470A	PPB	18 Jul 2023 12:36:24	U	SMPL
O3633-02 - 1	SS-01-COMP	188	0.0535	-	7470A	PPB	18 Jul 2023 12:38:40	U	SMPL
O3638-02 - 1	TP1	147	0.0447	-	7470A	PPB	18 Jul 2023 12:40:57	U	SMPL
O3638-12 - 1	TP2	227	0.0618	-	7470A	PPB	18 Jul 2023 12:43:13	U	SMPL
O3638-22 - 1	TP6	319	0.0816	-	7470A	PPB	18 Jul 2023 12:45:28	U	SMPL
CCV57 - 1	CCV57	22536	4.8459	-	7470A	PPB	18 Jul 2023 12:47:43	U	SMPL
CCB57 - 1	CCB57	-24	0.0080	-	7470A	PPB	18 Jul 2023 12:49:58	U	SMPL
O3638-32 - 1	TP7	347	0.0876	-	7470A	PPB	18 Jul 2023 12:52:13	U	SMPL
O3639-02 - 1	TP3	382	0.0951	-	7470A	PPB	18 Jul 2023 12:54:28	U	SMPL
O3639-12 - 1	TP4	355	0.0893	-	7470A	PPB	18 Jul 2023 12:56:44	U	SMPL
O3639-22 - 1	TP5	257	0.0683	-	7470A	PPB	18 Jul 2023 12:59:00	U	SMPL
O3639-32 - 1	TP10	212	0.0586	-	7470A	PPB	18 Jul 2023 13:01:15	U	SMPL
O3640-02 - 1	TP9	203	0.0567	-	7470A	PPB	18 Jul 2023 13:03:32	U	SMPL
O3640-12 - 1	TP8	184	0.0526	-	7470A	PPB	18 Jul 2023 13:05:48	U	SMPL
O3640-22 - 1	TP12	176	0.0509	-	7470A	PPB	18 Jul 2023 13:08:05	U	SMPL
O3640-32 - 1	TP11	128	0.0406	-	7470A	PPB	18 Jul 2023 13:10:22	U	SMPL
O3640-32DUP - 1	TP11DUP	87	0.0318	-	7470A	PPB	18 Jul 2023 13:12:38	U	SMPL
CCV58 - 1	CCV58	22058	4.7434	-	7470A	PPB	18 Jul 2023 13:14:53	U	SMPL
CCB58 - 1	CCB58	-24	0.0080	-	7470A	PPB	18 Jul 2023 13:17:08	U	SMPL
O3640-32MS - 1	TP11MS	16497	3.5508	-	7470A	PPB	18 Jul 2023 13:19:23	U	SMPL
O3640-32MSD - 1	TP11MSD	16852	3.6270	-	7470A	PPB	18 Jul 2023 13:21:38	U	SMPL
O3641-02 - 1	TP13	95	0.0335	-	7470A	PPB	18 Jul 2023 13:23:54	U	SMPL
O3641-12 - 1	TP14	61	0.0262	-	7470A	PPB	18 Jul 2023 13:26:10	U	SMPL
O3641-22 - 1	TP16	50	0.0239	-	7470A	PPB	18 Jul 2023 13:28:26	U	SMPL
O3641-32 - 1	TP15	74	0.0290	-	7470A	PPB	18 Jul 2023 13:30:42	U	SMPL
PB154248BL - 1	PBW	106	0.0359	-	7470A	PPB	18 Jul 2023 13:32:58	U	SMPL
PB154248BS - 1	LCSW	16130	3.4721	-	7470A	PPB	18 Jul 2023 13:35:15	U	SMPL
O3615-08 - 1	2880	4	0.0140	-	7470A	PPB	18 Jul 2023 13:37:33	U	SMPL
O3615-08DUP - 1	2880DUP	49	0.0237	-	7470A	PPB	18 Jul 2023 13:39:48	U	SMPL
CCV59 - 1	CCV59	22627	4.8654	-	7470A	PPB	18 Jul 2023 13:42:04	U	SMPL
CCB59 - 1	CCB59	-25	0.0078	-	7470A	PPB	18 Jul 2023 13:44:19	U	SMPL
O3615-08MS - 1	2880MS	16401	3.5303	-	7470A	PPB	18 Jul 2023 13:46:42	U	SMPL
O3615-08MSD - 1	2880MSD	16919	3.6413	-	7470A	PPB	18 Jul 2023 13:49:06	U	SMPL
PB154217TB - 1	PB154217TB	30	0.0196	-	7470A	PPB	18 Jul 2023 13:51:27	U	SMPL
PB154229TB - 1	PB154229TB	54	0.0247	-	7470A	PPB	18 Jul 2023 13:53:48	U	SMPL
O3640-32LX5 - 1		36	0.0209	-	7470A	PPB	18 Jul 2023 13:56:09	U	SMPL
O3640-32A - 1		17591	3.7854	-	7470A	PPB	18 Jul 2023 13:58:30	U	SMPL
O3615-08LX5 - 1		-50	0.0024	-	7470A	PPB	18 Jul 2023 14:00:51	U	SMPL
O3615-08A - 1		18288	3.9349	-	7470A	PPB	18 Jul 2023 14:03:12	U	SMPL
CCV60 - 1	CCV60	23088	4.9642	-	7470A	PPB	18 Jul 2023 14:05:33	U	SMPL
CCB60 - 1	CCB60	-55	0.0014	-	7470A	PPB	18 Jul 2023 14:07:54	U	SMPL