

LB126505


7470A

INSTRUMENT ID: CY1

Linear

 $\mu$  Abs.:

44037



A= 0.0000e+000

B= 2.3095e-004

C= -6.8958e-002

Rho= 0.9997971

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.037  | 0.037  | 458   | 0.000      | 458   | 0     |       |       |       | 1  |
| 0.2    | 0.200  | 0.206  | 0.006  | 1190  | 0.0 %      | 1190  |       |       |       |       | 3  |
| 2.5    | 2.500  | 2.549  | 0.049  | 11335 | 0.0 %      | 11335 |       |       |       |       | 2  |
| 5.0    | 5.000  | 4.889  | -0.111 | 21469 | 0.0 %      | 21469 |       |       |       |       | 2  |
| 7.5    | 7.500  | 7.417  | -0.083 | 32415 | 0.0 %      | 32415 |       |       |       |       | 1  |
| 10.0   | 10.000 | 10.102 | 0.102  | 44037 | 0.0 %      | 44037 |       |       |       |       | 1  |

## LB126505 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jul 2023 09:04:32

| Sample ID       | Extended ID   | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|---------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50            | 458        | -       | 0.0000   | 7470A  | PPB   | 19 Jul 2023 09:36:19 | S    | Std  |
| 0.2 - 1         | 50.2          | 1190       | -       | 0.2000   | 7470A  | PPB   | 19 Jul 2023 09:38:38 | S    | Std  |
| 2.5 - 1         | 52.5          | 11335      | -       | 2.5000   | 7470A  | PPB   | 19 Jul 2023 09:40:55 | S    | Std  |
| 5.0 - 1         | 55            | 21469      | -       | 5.0000   | 7470A  | PPB   | 19 Jul 2023 09:43:13 | S    | Std  |
| 7.5 - 1         | 57.5          | 32415      | -       | 7.5000   | 7470A  | PPB   | 19 Jul 2023 09:45:29 | S    | Std  |
| 10.0 - 1        | 510           | 44037      | -       | 10.0000  | 7470A  | PPB   | 19 Jul 2023 09:47:44 | S    | Std  |
| ICV75 - 1       | ICV75         | 18637      | 4.2353  | -        | 7470A  | PPB   | 19 Jul 2023 09:50:39 | U    | SMPL |
| ICB75 - 1       | ICB75         | 156        | -0.0329 | -        | 7470A  | PPB   | 19 Jul 2023 09:52:54 | U    | SMPL |
| CCV61 - 1       | CCV61         | 21372      | 4.8670  | -        | 7470A  | PPB   | 19 Jul 2023 09:55:08 | U    | SMPL |
| CCB61 - 1       | CCB61         | 183        | -0.0267 | -        | 7470A  | PPB   | 19 Jul 2023 09:57:24 | U    | SMPL |
| CRA - 1         | CRA           | 1145       | 0.1955  | -        | 7470A  | PPB   | 19 Jul 2023 09:59:40 | U    | SMPL |
| HighStd - 1     | HighStd       | 44434      | 10.1933 | -        | 7470A  | PPB   | 19 Jul 2023 10:01:55 | U    | SMPL |
| ChkStd - 1      | ChkStd        | 30652      | 7.0103  | -        | 7470A  | PPB   | 19 Jul 2023 10:04:10 | U    | SMPL |
| PB154279BL - 1  | PBW           | 409        | 0.0255  | -        | 7470A  | PPB   | 19 Jul 2023 10:06:26 | U    | SMPL |
| PB154279BS - 1  | LCSW          | 15791      | 3.5780  | -        | 7470A  | PPB   | 19 Jul 2023 10:08:42 | U    | SMPL |
| O3637-01 - 1    | A508          | 12967      | 2.9258  | -        | 7470A  | PPB   | 19 Jul 2023 10:11:00 | U    | SMPL |
| O3637-01DUP - 1 | A508DUP       | 11053      | 2.4838  | -        | 7470A  | PPB   | 19 Jul 2023 10:13:16 | U    | SMPL |
| O3637-01MS - 1  | A508MS        | 27870      | 6.3677  | -        | 7470A  | PPB   | 19 Jul 2023 10:15:39 | U    | SMPL |
| O3637-01MSD - 1 | A508MSD       | 27991      | 6.3957  | -        | 7470A  | PPB   | 19 Jul 2023 10:17:59 | U    | SMPL |
| O3645-08 - 1    | RINSATE-BLANK | 273        | -0.0059 | -        | 7470A  | PPB   | 19 Jul 2023 10:20:15 | U    | SMPL |
| CCV62 - 1       | CCV62         | 21403      | 4.8742  | -        | 7470A  | PPB   | 19 Jul 2023 10:22:32 | U    | SMPL |
| CCB62 - 1       | CCB62         | 237        | -0.0142 | -        | 7470A  | PPB   | 19 Jul 2023 10:24:49 | U    | SMPL |
| O3637-01LX5 - 1 |               | 2369       | 0.4782  | -        | 7470A  | PPB   | 19 Jul 2023 10:27:04 | U    | SMPL |
| O3637-01A - 1   |               | 29083      | 6.6479  | -        | 7470A  | PPB   | 19 Jul 2023 10:29:19 | U    | SMPL |
| CCV63 - 1       | CCV63         | 21530      | 4.9035  | -        | 7470A  | PPB   | 19 Jul 2023 10:31:34 | U    | SMPL |
| CCB63 - 1       | CCB63         | 217        | -0.0188 | -        | 7470A  | PPB   | 19 Jul 2023 10:33:51 | U    | SMPL |