



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126508

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/19/2023	08:30
2	CAL2	1	Water	NA	NA	20.7	7.00	07/19/2023	08:31
3	CAL3	1	Water	NA	NA	20.8	10.03	07/19/2023	08:33
4	ICV	1	Water	NA	NA	20.6	6.98	07/19/2023	08:35
5	CCV1	1	Water	NA	NA	20.6	2.01	07/19/2023	08:36
6	O3651-01	1	Water	NA	NA	20.2	7.16	07/19/2023	08:40
7	O3651-01DUP	1	Water	NA	NA	20.3	7.17	07/19/2023	08:41
8	O3664-03	1	Water	NA	NA	20.1	6.02	07/19/2023	08:45
9	O3664-04	1	Water	NA	NA	20.2	6.16	07/19/2023	08:46
10	O3664-05	1	Water	NA	NA	20.5	4.84	07/19/2023	08:50
11	O3664-06	1	Water	NA	NA	20.6	5.28	07/19/2023	09:00
12	O3664-07	1	Water	NA	NA	20.4	6.62	07/19/2023	09:02
13	CCV2	1	Water	NA	NA	20.8	12.02	07/19/2023	09:10

WORKLIST(Hardcopy Internal Chain)

WB 26508

WorkList Name : ph 03651

WorkList ID : 171985

Department : Wet-Chemistry

Date : 07-19-2023 07:55:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3651-01	NWB-1925	Water	pH	Cool 4 deg C	PSEG03	I11	07/17/2023	9040C
O3664-03	AUD-1007	Water	pH	Cool 4 deg C	PSEG03	I11	07/18/2023	9040C
O3664-04	AUD-1013	Water	pH	Cool 4 deg C	PSEG03	I11	07/18/2023	9040C
O3664-05	AUD-1012	Water	pH	Cool 4 deg C	PSEG03	I11	07/18/2023	9040C
O3664-06	AUD-1022	Water	pH	Cool 4 deg C	PSEG03	I11	07/18/2023	9040C
O3664-07	A462	Water	pH	Cool 4 deg C	PSEG03	I11	07/18/2023	9040C

Date/Time 07/19/23 08:00

Raw Sample Received by: JDL

Raw Sample Relinquished by: JDL

Date/Time 07/19/23

Raw Sample Received by: JDL

Raw Sample Relinquished by: JDL

12:30