



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126509

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	07/19/2023	08:30
2	CAL2	1	Water	NA	NA	20.7	7.00	07/19/2023	08:31
3	CAL3	1	Water	NA	NA	20.8	10.03	07/19/2023	08:33
4	ICV	1	Water	NA	NA	20.6	6.98	07/19/2023	08:35
5	CCV1	1	Water	NA	NA	20.9	2.01	07/19/2023	10:25
6	O3664-01	1	Solid	20.02	20	20.7	5.99	07/19/2023	10:30
7	O3664-01DUP	1	Solid	20.01	20	20.8	6.00	07/19/2023	10:31
8	O3664-02	1	Solid	20.03	20	20.6	6.01	07/19/2023	10:37
9	CCV2	1	Water	NA	NA	20.8	12.02	07/19/2023	10:40

283126509

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O3664 S

WorkList ID : 171995

Department : Wet-Chemistry

Date : 07-19-2023 10:12:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3664-01	AUD-1021-1020-COMP	Solid	pH	Cool 4 deg C	PSEG03	I11	07/18/2023	9045D
O3664-02	AUD-1023	Solid	pH	Cool 4 deg C	PSEG03	I11	07/18/2023	9045D

Date/Time 07-19-23 10:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 07-19-23 13:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]