



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB126521

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/19/2023	17:00
2	CAL2	1	Water	NA	NA	20.8	7.00	07/19/2023	17:01
3	CAL3	1	Water	NA	NA	20.7	10.02	07/19/2023	17:04
4	ICV	1	Water	NA	NA	20.7	6.98	07/19/2023	17:05
5	CCV1	1	Water	NA	NA	20.7	2.01	07/19/2023	17:35
6	O3668-05	1	Water	NA	NA	22.5	7.20	07/19/2023	17:40
7	O3668-05DUP	1	Water	NA	NA	22.6	7.21	07/19/2023	17:41
8	O3682-04	1	Water	NA	NA	23.4	7.79	07/19/2023	17:45
9	CCV2	1	Water	NA	NA	20.8	12.02	07/19/2023	17:49

VB126521

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 03682 W WorkList ID : 172008 Department : Wet-Chemistry Date : 07-19-2023 14:55:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3668-05	N48719	Water	Corrosivity	Cool 4 deg C	PSEG03	I11	07/19/2023	9040C
O3682-04	ELZ-23-00059	Water	Corrosivity	Cool 4 deg C	PSEG03	I11	07/19/2023	9040C

Date/Time 07-19-23 16:50
 Raw Sample Received by: JBC
 Raw Sample Relinquished by: JBC

Date/Time 07-19-23 19:00
 Raw Sample Received by: JBC
 Raw Sample Relinquished by: JBC