



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126534

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/20/2023	11:25
2	CAL2	1	Water	NA	NA	20.7	7.00	07/20/2023	11:26
3	CAL3	1	Water	NA	NA	20.6	10.02	07/20/2023	11:30
4	ICV	1	Water	NA	NA	20.7	6.99	07/20/2023	11:35
5	CCV1	1	Water	NA	NA	20.8	2.01	07/20/2023	11:40
6	O3677-01	1	Solid	20.02	20	22.6	6.22	07/20/2023	11:45
7	O3681-01	1	Solid	20.03	20	22.5	8.56	07/20/2023	11:47
8	O3681-04	1	Solid	20.04	20	22.5	8.96	07/20/2023	11:50
9	O3681-07	1	Solid	20.02	20	22.3	9.28	07/20/2023	12:00
10	O3681-10	1	Solid	20.03	20	22.5	9.13	07/20/2023	12:10
11	O3681-13	1	Solid	20.05	20	22.3	8.78	07/20/2023	12:11
12	O3681-16	1	Solid	20.05	20	22.5	8.93	07/20/2023	12:15
13	O3681-19	1	Solid	20.02	20	22.3	9.15	07/20/2023	12:20
14	O3681-22	1	Solid	20.01	20	22.4	8.98	07/20/2023	12:25
15	O3681-22DUP	1	Solid	20.02	20	22.5	8.99	07/20/2023	12:26
16	CCV2	1	Water	NA	NA	20.8	12.03	07/20/2023	12:30

126534

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 03681 WorkList ID : 172039 Department : Wet-Chemistry Date : 07-20-2023 11:13:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3677-01	SWC-1	Solid	Corrosivity	Cool 4 deg C	REMI02	I11	07/18/2023	9045D
O3681-01	RPTP-1-CS-1	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D
O3681-04	RPTP-2-CS-2	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D
O3681-07	RPTP-3-CS-3	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D
O3681-10	RPTP-4-CS-4	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D
O3681-13	RPTP-5-CS-5	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D
O3681-16	RPTP-6-CS-6	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D
O3681-19	RPTP-7-CS-7	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D
O3681-22	RPTP-8-CS-8	Solid	Corrosivity	Cool 4 deg C	PREF02	I11	07/17/2023	9045D

Date/Time07/20/2311:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time07/20/2313:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]