

LB126541


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

66575



A= 0.0000e+000

B= 1.5071e-004

C= 3.8839e-002

Rho= 0.9997552

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.066	0.066	178	0.000	178					-
0.2	0.200	0.165	-0.035	836	0.0 %	836					-18
2.5	2.500	2.455	-0.045	16031	0.0 %	16031					-2
5.0	5.000	5.084	0.084	33478	0.0 %	33478					2
7.5	7.500	7.358	-0.142	48567	0.0 %	48567					-2
10.0	10.000	10.072	0.072	66575	0.0 %	66575					1

LB126541 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jul 2023 08:36:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	178	-	0.0000	7471B	PPB	21 Jul 2023 08:48:07	S	Std
0.2 - 1	50.2	836	-	0.2000	7471B	PPB	21 Jul 2023 08:50:23	S	Std
2.5 - 1	52.5	16031	-	2.5000	7471B	PPB	21 Jul 2023 08:52:40	S	Std
5.0 - 1	55	33478	-	5.0000	7471B	PPB	21 Jul 2023 08:54:57	S	Std
7.5 - 1	55.5	48567	-	7.5000	7471B	PPB	21 Jul 2023 08:57:13	S	Std
10.0 - 1	511	66575	-	10.0000	7471B	PPB	21 Jul 2023 08:59:29	S	Std
ICV79 - 1	ICV79	26337	4.0080	-	7471B	PPB	21 Jul 2023 09:02:15	U	SMPL
ICB79 - 1	ICB79	0	0.0388	-	7471B	PPB	21 Jul 2023 09:04:30	U	SMPL
CCV70 - 1	CCV70	30069	4.5705	-	7471B	PPB	21 Jul 2023 09:06:46	U	SMPL
CCB70 - 1	CCB70	-15	0.0366	-	7471B	PPB	21 Jul 2023 09:09:00	U	SMPL
CRA - 1	CRA	870	0.1700	-	7471B	PPB	21 Jul 2023 09:15:59	U	SMPL
HighStd - 1	HighStd	65474	9.9063	-	7471B	PPB	21 Jul 2023 09:18:14	U	SMPL
ChkStd - 1	ChkStd	45323	6.8694	-	7471B	PPB	21 Jul 2023 09:20:30	U	SMPL
PB154339BL - 1	PBS	-252	0.0009	-	7471B	PPB	21 Jul 2023 09:22:48	U	SMPL
PB154339BS - 1	LCSS	23637	3.6011	-	7471B	PPB	21 Jul 2023 09:25:06	U	SMPL
O3655-01 - 1	MARION-COMP	33419	5.0753	-	7471B	PPB	21 Jul 2023 09:27:23	U	SMPL
O3657-01 - 1	BERGENSWITCH-COMP1	1669	0.2904	-	7471B	PPB	21 Jul 2023 09:29:41	U	SMPL
O3657-03 - 1	BERGENSWITCH-COMP2	1028	0.1938	-	7471B	PPB	21 Jul 2023 09:32:01	U	SMPL
O3659-01 - 1	HR-2-071823	2671	0.4414	-	7471B	PPB	21 Jul 2023 09:34:18	U	SMPL
O3659-01DUP - 1	HR-2-071823DUP	3013	0.4929	-	7471B	PPB	21 Jul 2023 09:36:33	U	SMPL
CCV71 - 1	CCV71	29995	4.5593	-	7471B	PPB	21 Jul 2023 09:38:48	U	SMPL
CCB71 - 1	CCB71	-17	0.0363	-	7471B	PPB	21 Jul 2023 09:41:03	U	SMPL
O3659-01MS - 1	HR-2-071823MS	28010	4.2602	-	7471B	PPB	21 Jul 2023 09:43:20	U	SMPL
O3659-01MSD - 1	HR-2-071823MSD	28742	4.3705	-	7471B	PPB	21 Jul 2023 09:45:36	U	SMPL
O3659-03 - 1	HR-3-071823	5384	0.8502	-	7471B	PPB	21 Jul 2023 09:47:54	U	SMPL
O3659-05 - 1	HR-4-071823	4260	0.6809	-	7471B	PPB	21 Jul 2023 09:50:11	U	SMPL
O3662-01 - 1	BUR-1276	2412	0.4023	-	7471B	PPB	21 Jul 2023 09:52:27	U	SMPL
O3664-01 - 1	AUD-1021-1020-COMP	807	0.1605	-	7471B	PPB	21 Jul 2023 09:54:43	U	SMPL
O3664-02 - 1	AUD-1023	1374	0.2459	-	7471B	PPB	21 Jul 2023 09:57:00	U	SMPL
O3668-01 - 1	SLUDGE-DRUM	10233	1.5810	-	7471B	PPB	21 Jul 2023 09:59:17	U	SMPL
O3668-03 - 1	332	8416	1.3072	-	7471B	PPB	21 Jul 2023 10:01:33	U	SMPL
O3669-01 - 1	DEL-AVE-1	466	0.1091	-	7471B	PPB	21 Jul 2023 10:03:51	U	SMPL
CCV72 - 1	CCV72	30373	4.6163	-	7471B	PPB	21 Jul 2023 10:06:06	U	SMPL
CCB72 - 1	CCB72	-39	0.0330	-	7471B	PPB	21 Jul 2023 10:08:22	U	SMPL
O3669-02 - 1	DEL-AVE-2	612	0.1311	-	7471B	PPB	21 Jul 2023 10:10:40	U	SMPL
O3678-01 - 1	LWC-1	138	0.0596	-	7471B	PPB	21 Jul 2023 10:12:56	U	SMPL
O3678-02 - 1	LWC-2	122	0.0572	-	7471B	PPB	21 Jul 2023 10:15:11	U	SMPL
O3678-03 - 1	LWC-3	275	0.0803	-	7471B	PPB	21 Jul 2023 10:17:27	U	SMPL
PB154340BL - 1	PBS	165	0.0637	-	7471B	PPB	21 Jul 2023 10:19:43	U	SMPL
PB154340BS - 1	LCSS	24230	3.6905	-	7471B	PPB	21 Jul 2023 10:24:41	U	SMPL
O3680-01 - 1	TP-3	764	0.1540	-	7471B	PPB	21 Jul 2023 10:26:57	U	SMPL
O3681-01 - 1	RPTP-1-CS-1	4995	0.7916	-	7471B	PPB	21 Jul 2023 10:29:14	U	SMPL
O3681-04 - 1	RPTP-2-CS-2	10050	1.5534	-	7471B	PPB	21 Jul 2023 10:31:31	U	SMPL
O3681-07 - 1	RPTP-3-CS-3	5776	0.9093	-	7471B	PPB	21 Jul 2023 10:33:46	U	SMPL
CCV73 - 1	CCV73	33876	5.1442	-	7471B	PPB	21 Jul 2023 10:36:01	U	SMPL
CCB73 - 1	CCB73	39	0.0447	-	7471B	PPB	21 Jul 2023 10:38:17	U	SMPL
O3681-10 - 1	RPTP-4-CS-4	5555	0.8760	-	7471B	PPB	21 Jul 2023 10:40:32	U	SMPL
O3681-13 - 1	RPTP-5-CS-5	7271	1.1346	-	7471B	PPB	21 Jul 2023 10:42:47	U	SMPL
O3681-16 - 1	RPTP-6-CS-6	6676	1.0450	-	7471B	PPB	21 Jul 2023 10:45:03	U	SMPL
O3681-19 - 1	RPTP-7-CS-7	6016	0.9455	-	7471B	PPB	21 Jul 2023 10:47:19	U	SMPL
O3681-22 - 1	RPTP-8-CS-8	1105	0.2054	-	7471B	PPB	21 Jul 2023 10:49:34	U	SMPL
O3685-01 - 1	CHRT-24152	8199	1.2745	-	7471B	PPB	21 Jul 2023 11:02:17	U	SMPL
O3685-01DUP - 1	CHRT-24152DUP	8994	1.3943	-	7471B	PPB	21 Jul 2023 11:04:33	U	SMPL
O3685-01MS - 1	CHRT-24152MS	31599	4.8010	-	7471B	PPB	21 Jul 2023 11:06:50	U	SMPL
O3685-01MSD - 1	CHRT-24152MSD	30513	4.6374	-	7471B	PPB	21 Jul 2023 11:11:44	U	SMPL
O3686-01 - 1	SB-21-0-2-20230719	21134	3.2239	-	7471B	PPB	21 Jul 2023 11:14:00	U	SMPL
CCV74 - 1	CCV74	33511	5.0892	-	7471B	PPB	21 Jul 2023 11:16:15	U	SMPL
CCB74 - 1	CCB74	40	0.0449	-	7471B	PPB	21 Jul 2023 11:18:31	U	SMPL
O3686-03 - 1	SB-26-0-2-20230719	8327	1.2938	-	7471B	PPB	21 Jul 2023 11:20:46	U	SMPL
O3694-01 - 1	WC-1	1662	0.2893	-	7471B	PPB	21 Jul 2023 11:23:02	U	SMPL
O3659-01LX5 - 1		392	0.0979	-	7471B	PPB	21 Jul 2023 11:25:17	U	SMPL
O3659-01A - 1		27898	4.2433	-	7471B	PPB	21 Jul 2023 11:27:32	U	SMPL
O3685-01LX5 - 1		1272	0.2305	-	7471B	PPB	21 Jul 2023 11:29:48	U	SMPL
O3685-01A - 1		33353	5.0654	-	7471B	PPB	21 Jul 2023 11:32:05	U	SMPL
CCV75 - 1	CCV75	33591	5.1013	-	7471B	PPB	21 Jul 2023 11:34:22	U	SMPL
CCB75 - 1	CCB75	-68	0.0286	-	7471B	PPB	21 Jul 2023 11:36:39	U	SMPL