



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126549

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP103373
Residual chlorine ICV-LCS, 0.4PPM	WP103372
Chlorine Calibration std, 0.1ppm	WP103367
Chlorine Calibration std, 0.2ppm	WP103368
Chlorine Calibration std, 0.8ppm	WP103370
Chlorine Calibration std, 0.0ppm	WP103366
Chlorine Calibration std, 1.6ppm	WP103371
Residual Chlorine LOQ 0.1ppm	WP103374
Residual Chlorine Calibration and CCV std, 0	WP103369
Total Chlorine Powder Pillows	W2759

Intercept: 0.0080

Slope: 0.9974

Regression: 0.999800

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		07/21/2023	10:20
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	12	07/21/2023	10:23
3	CAL3	0.2	1	0.000	0.210	0.210	0.20	1.5	07/21/2023	10:26
4	CAL4	0.4	1	0.000	0.390	0.390	0.38	-4.3	07/21/2023	10:30
5	CAL5	0.8	1	0.000	0.820	0.820	0.81	1.7	07/21/2023	10:33
6	CAL6	1.6	1	0.000	1.600	1.600	1.60	-0.3	07/21/2023	10:36



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126549

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4030	07/21/2023	10:40
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	07/21/2023	10:43
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.3930	07/21/2023	10:46
4	CCB1				1	0.0000	0.0100	0.0100	0.0020	07/21/2023	10:50
5	LB126549BL	50	50		1	0.0000	0.0100	0.0100	0.0020	07/21/2023	10:53
6	LB126549BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.4130	07/21/2023	10:59
7	O3572-07	50	50		1	0.0000	0.0500	0.0500	0.0420	07/21/2023	11:02
8	O3572-08	50	50		1	0.0000	0.1200	0.1200	0.1120	07/21/2023	11:05
9	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4030	07/21/2023	11:08
10	CCB2				1	0.0000	0.0000	0.0000	-0.0080	07/21/2023	11:12