



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126550

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	07/21/2023	08:45
2	O3668-05		1	12	100.00	>212.0	>212.0	07/21/2023	09:15
3	O3670-01		1	11	100.00	>212.0	>212.0	07/21/2023	09:45
4	O3682-04		1	14	100.00	>212.0	>212.0	07/21/2023	10:15
5	O3683-01		1	11	100.00	>212.0	>212.0	07/21/2023	10:45
6	O3688-01		1	12	100.00	>212.0	>212.0	07/21/2023	11:15
7	O3688-01DUP		1	12	100.00	>212.0	>212.0	07/21/2023	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6126550

WorkList Name : fp 7-20 WorkList ID : 172029 Department : Wet-Chemistry Date : 07-20-2023 08:29:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3668-05 F	N48719	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/19/2023	1010B
O3670-01 J	N48909	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/19/2023	1010B
O3682-04 D	ELZ-23-00059	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/19/2023	1010B
O3683-01 L	NORTH-TURNPIKE	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/19/2023	1010B
O3688-01 D	FMI-123	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/19/2023	1010B

Date/Time 07/21/2023 08:40
 Raw Sample Received by: RM awc
 Raw Sample Relinquished by: 78(awc)

Date/Time 07/21/2023 12:20
 Raw Sample Received by: 78(awc)
 Raw Sample Relinquished by: RM awc