

LB126559


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

60570



A= 0.0000e+000

B= 1.6622e-004

C= -8.1761e-002

Rho= 0.9997443

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.045	-0.045	224	0.000	224					
0.2	0.200	0.171	-0.029	1519	0.0 %	1519					
2.5	2.500	2.529	0.029	15708	0.0 %	15708					
5.0	5.000	5.164	0.164	31556	0.0 %	31556					
7.5	7.500	7.394	-0.106	44977	0.0 %	44977					
10.0	10.000	9.986	-0.014	60570	0.0 %	60570					

0.0
-15
-13
-10

LB126559 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jul 2023 08:29:03

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	224	-	0.00007471B	PPB		24 Jul 2023 08:43:17	S	Std
0.2 - 1	50.2	1519	-	0.20007471B	PPB		24 Jul 2023 08:45:34	S	Std
2.5 - 1	52.5	15708	-	2.50007471B	PPB		24 Jul 2023 08:47:51	S	Std
5.0 - 1	55	31556	-	5.00007471B	PPB		24 Jul 2023 08:50:08	S	Std
7.5 - 1	57.5	44977	-	7.50007471B	PPB		24 Jul 2023 08:55:00	S	Std
10.0 - 1	510	60570	-	10.00007471B	PPB		24 Jul 2023 08:59:51	S	Std
ICV81 - 1	ICV81	23220	3.7780	-7471B	PPB		24 Jul 2023 09:02:46	U	SMPL
ICB81 - 1	ICB81	80	-0.0685	-7471B	PPB		24 Jul 2023 09:05:03	U	SMPL
CCV79 - 1	CCV79	29384	4.8026	-7471B	PPB		24 Jul 2023 09:07:17	U	SMPL
CCB79 - 1	CCB79	67	-0.0706	-7471B	PPB		24 Jul 2023 09:09:32	U	SMPL
CRA - 1	CRA	1461	0.1611	-7471B	PPB		24 Jul 2023 09:11:48	U	SMPL
HighStd - 1	HighStd	60304	9.9422	-7471B	PPB		24 Jul 2023 09:14:02	U	SMPL
ChkStd - 1	ChkStd	42149	6.9244	-7471B	PPB		24 Jul 2023 09:16:17	U	SMPL
PB154359BL - 1	PBS	-51	-0.0902	-7471B	PPB		24 Jul 2023 09:18:33	U	SMPL
PB154359BS - 1	LCSS	29185	4.7695	-7471B	PPB		24 Jul 2023 09:20:50	U	SMPL
O3686-05 - 1	SB-20-0-2-20230720	15853	2.5534	-7471B	PPB		24 Jul 2023 09:23:07	U	SMPL
O3686-07 - 1	SB-18-0-2-20230720	21012	3.4109	-7471B	PPB		24 Jul 2023 09:25:22	U	SMPL
O3706-01 - 1	B-P9A	1510	0.1692	-7471B	PPB		24 Jul 2023 09:27:40	U	SMPL
O3706-02 - 1	B-P9B	670	0.0296	-7471B	PPB		24 Jul 2023 09:29:57	U	SMPL
O3706-03 - 1	B-P10A	1976	0.2467	-7471B	PPB		24 Jul 2023 09:32:12	U	SMPL
CCV80 - 1	CCV80	29375	4.8011	-7471B	PPB		24 Jul 2023 09:34:27	U	SMPL
CCB80 - 1	CCB80	56	-0.0725	-7471B	PPB		24 Jul 2023 09:36:42	U	SMPL
O3706-04 - 1	B-P10B	641	0.0248	-7471B	PPB		24 Jul 2023 09:38:57	U	SMPL
O3706-05 - 1	B-P20A	1415	0.1534	-7471B	PPB		24 Jul 2023 09:41:12	U	SMPL
O3706-06 - 1	B-P20B	734	0.0402	-7471B	PPB		24 Jul 2023 09:43:28	U	SMPL
O3706-07 - 1	B-P21A	1469	0.1624	-7471B	PPB		24 Jul 2023 09:45:43	U	SMPL
O3706-08 - 1	B-P21B	1508	0.1689	-7471B	PPB		24 Jul 2023 09:47:59	U	SMPL
O3707-01 - 1	EO-01-072023	3681	0.5301	-7471B	PPB		24 Jul 2023 09:50:14	U	SMPL
O3708-01 - 1	PL-02-072023	1608	0.1855	-7471B	PPB		24 Jul 2023 09:52:30	U	SMPL
O3717-01 - 1	TP-1	707	0.0358	-7471B	PPB		24 Jul 2023 09:54:47	U	SMPL
O3717-04 - 1	TP-2	1230	0.1227	-7471B	PPB		24 Jul 2023 09:57:04	U	SMPL
O3717-04DUP - 1	TP-2DUP	1411	0.1528	-7471B	PPB		24 Jul 2023 09:59:19	U	SMPL
CCV81 - 1	CCV81	29400	4.8052	-7471B	PPB		24 Jul 2023 10:01:35	U	SMPL
CCB81 - 1	CCB81	52	-0.0731	-7471B	PPB		24 Jul 2023 10:03:50	U	SMPL
O3717-04MS - 1	TP-2MS	28507	4.6568	-7471B	PPB		24 Jul 2023 10:06:06	U	SMPL
O3717-04MSD - 1	TP-2MSD	28683	4.6860	-7471B	PPB		24 Jul 2023 10:08:21	U	SMPL
O3723-01 - 1	S-1	21329	3.4636	-7471B	PPB		24 Jul 2023 10:10:38	U	SMPL
O3723-02 - 1	S-2	1679	0.1973	-7471B	PPB		24 Jul 2023 10:12:54	U	SMPL
O3723-03 - 1	S-3	6264	0.9595	-7471B	PPB		24 Jul 2023 10:15:10	U	SMPL
PB154360BL - 1	PBS	68	-0.0705	-7471B	PPB		24 Jul 2023 10:17:26	U	SMPL
PB154360BS - 1	LCSS	21186	3.4399	-7471B	PPB		24 Jul 2023 10:19:42	U	SMPL
O3715-01 - 1	T842-FH-01-CT	506	0.0023	-7471B	PPB		24 Jul 2023 10:21:59	U	SMPL
O3715-02 - 1	T842-FH-01-BW	1028	0.0891	-7471B	PPB		24 Jul 2023 10:24:15	U	SMPL
O3715-03 - 1	T842-FH-02-CT	216	-0.0459	-7471B	PPB		24 Jul 2023 10:26:31	U	SMPL
CCV82 - 1	CCV82	29473	4.8174	-7471B	PPB		24 Jul 2023 10:28:46	U	SMPL
CCB82 - 1	CCB82	49	-0.0736	-7471B	PPB		24 Jul 2023 10:31:01	U	SMPL
O3715-04 - 1	T842-FH-02-BW	334	-0.0262	-7471B	PPB		24 Jul 2023 10:33:17	U	SMPL
O3715-05 - 1	T842-FH-03-CT	234	-0.0429	-7471B	PPB		24 Jul 2023 10:35:34	U	SMPL
O3715-06 - 1	T842-FH-03-BW	248	-0.0405	-7471B	PPB		24 Jul 2023 10:37:50	U	SMPL
O3715-07 - 1	T842-FH-04-CT	308	-0.0306	-7471B	PPB		24 Jul 2023 10:40:05	U	SMPL
O3715-08 - 1	T842-FH-04-BW	286	-0.0342	-7471B	PPB		24 Jul 2023 10:42:21	U	SMPL
O3715-09 - 1	T842-CT-01	329	-0.0271	-7471B	PPB		24 Jul 2023 10:44:38	U	SMPL
O3715-10 - 1	T842-CT-02	695	0.0338	-7471B	PPB		24 Jul 2023 10:46:54	U	SMPL
O3715-11 - 1	T843-CT-01	368	-0.0206	-7471B	PPB		24 Jul 2023 10:49:12	U	SMPL
O3715-12 - 1	T841-BTW-01	290	-0.0336	-7471B	PPB		24 Jul 2023 10:51:30	U	SMPL
O3715-13 - 1	FIELD-BLANK	146	-0.0575	-7471B	PPB		24 Jul 2023 10:53:46	U	SMPL
CCV83 - 1	CCV83	31824	5.2081	-7471B	PPB		24 Jul 2023 10:56:02	U	SMPL
CCB83 - 1	CCB83	49	-0.0736	-7471B	PPB		24 Jul 2023 10:58:18	U	SMPL
O3719-01 - 1	B-P11A	2506	0.3348	-7471B	PPB		24 Jul 2023 11:00:33	U	SMPL
O3719-02 - 1	B-P11B	348	-0.0239	-7471B	PPB		24 Jul 2023 11:02:49	U	SMPL
O3719-03 - 1	B-P22A	2542	0.3408	-7471B	PPB		24 Jul 2023 11:05:03	U	SMPL
O3719-04 - 1	B-P22B	1107	0.1022	-7471B	PPB		24 Jul 2023 11:07:19	U	SMPL
O3719-05 - 1	B-P23A	1696	0.2002	-7471B	PPB		24 Jul 2023 11:09:34	U	SMPL
O3719-06 - 1	B-P23B	302	-0.0316	-7471B	PPB		24 Jul 2023 11:11:50	U	SMPL
O3719-06DUP - 1	B-P23BDUP	286	-0.0342	-7471B	PPB		24 Jul 2023 11:14:07	U	SMPL
O3719-06MS - 1	B-P23BMS	26174	4.2690	-7471B	PPB		24 Jul 2023 11:16:18	U	SMPL
O3719-06MSD - 1	B-P23BMSD	24360	3.9675	-7471B	PPB		24 Jul 2023 11:18:35	U	SMPL
CCV84 - 1	CCV84	32005	5.2382	-7471B	PPB		24 Jul 2023 11:20:50	U	SMPL
CCB84 - 1	CCB84	40	-0.0751	-7471B	PPB		24 Jul 2023 11:23:06	U	SMPL
O3717-04LX5 - 1		252	-0.0399	-7471B	PPB		24 Jul 2023 11:25:22	U	SMPL
O3717-04A - 1		28993	4.7376	-7471B	PPB		24 Jul 2023 11:27:38	U	SMPL
O3719-06LX5 - 1		-36	-0.0877	-7471B	PPB		24 Jul 2023 11:29:54	U	SMPL
O3719-06A - 1		25005	4.0747	-7471B	PPB		24 Jul 2023 11:32:10	U	SMPL
CCV85 - 1	CCV85	30541	4.9949	-7471B	PPB		24 Jul 2023 11:34:25	U	SMPL
CCB85 - 1	CCB85	86	-0.0675	-7471B	PPB		24 Jul 2023 11:36:41	U	SMPL