



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126560

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/24/2023	08:10
2	CAL2	1	Water	NA	NA	20.7	7.00	07/24/2023	08:11
3	CAL3	1	Water	NA	NA	20.7	10.02	07/24/2023	08:12
4	ICV	1	Water	NA	NA	20.7	7.02	07/24/2023	08:15
5	CCV1	1	Water	NA	NA	20.7	2.01	07/24/2023	08:20
6	O3722-01	1	Solid	20.02	20	23.4	6.60	07/24/2023	08:25
7	O3722-01DUP	1	Solid	20.02	20	23.5	6.61	07/24/2023	08:26
8	O3722-02	1	Solid	20.01	20	23.3	6.39	07/24/2023	08:30
9	CCV2	1	Water	NA	NA	20.7	12.02	07/24/2023	08:33

WORKLIST(Hardcopy Internal Chain)

126560

WorkList Name : ph 03722 s WorkList ID : 172107 Department : Wet-Chemistry Date : 07-24-2023 07:56:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3722-01	34339	Solid	pH	Cool 4 deg C	PSEG03	J11	07/21/2023	9045D
O3722-02	34340	Solid	pH	Cool 4 deg C	PSEG03	J11	07/21/2023	9045D

Date/Time 07/24/23 08:00
 Raw Sample Received by: JD/WC
 Raw Sample Relinquished by: CD

Date/Time 07/24/23 11:00
 Raw Sample Received by: CD
 Raw Sample Relinquished by: SP/WC