


LB126567

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%d
0.0	0.000	-0.060	-0.060	45	0.000	45	0				
0.2	0.200	0.205	0.005	1171	0.0 %	1171					
2.5	2.500	2.621	0.121	11449	0.0 %	11449					
5.0	5.000	5.099	0.099	21988	0.0 %	21988					
7.5	7.500	7.210	-0.290	30967	0.0 %	30967					
10.0	10.000	10.125	0.125	43367	0.0 %	43367					

0/01
1
2
3
4
5

LB126567 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Jul 2023 12:17:57

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.000	7470A	PPB	24 Jul 2023 12:21:41	S	Std
0.2 - 1	50.2	1171	-	0.200	7470A	PPB	24 Jul 2023 12:23:59	S	Std
2.5 - 1	52.5	11449	-	2.500	7470A	PPB	24 Jul 2023 12:26:15	S	Std
5.0 - 1	55	21988	-	5.000	7470A	PPB	24 Jul 2023 12:31:07	S	Std
7.5 - 1	57.5	30967	-	7.500	7470A	PPB	24 Jul 2023 12:35:58	S	Std
10.0 - 1	51.5	43367	-	10.000	7470A	PPB	24 Jul 2023 12:41:56	S	Std
ICV82 - 1	ICV82	18265	4.2236	-	7470A	PPB	24 Jul 2023 12:49:34	U	SMPL
ICB82 - 1	ICB82	-44	-0.0807	-	7470A	PPB	24 Jul 2023 12:51:48	U	SMPL
CCV86 - 1	CCV86	22610	5.2451	-	7470A	PPB	24 Jul 2023 12:54:03	U	SMPL
CCB86 - 1	CCB86	-54	-0.0831	-	7470A	PPB	24 Jul 2023 12:56:18	U	SMPL
CRA - 1	CRA	1063	0.1795	-	7470A	PPB	24 Jul 2023 12:58:32	U	SMPL
HighStd - 1	HighStd	43509	10.1584	-	7470A	PPB	24 Jul 2023 13:00:47	U	SMPL
ChkStd - 1	ChkStd	29428	6.8480	-	7470A	PPB	24 Jul 2023 13:03:02	U	SMPL
PB154368BL - 1	PBW	-94	-0.0925	-	7470A	PPB	24 Jul 2023 13:05:19	U	SMPL
PB154368BS - 1	LCSW	14641	3.3716	-	7470A	PPB	24 Jul 2023 13:07:34	U	SMPL
O3688-02 - 1	SLUDGE-DRUM	958	0.1548	-	7470A	PPB	24 Jul 2023 13:09:51	U	SMPL
O3677-01 - 1	SWC-1	-47	-0.0814	-	7470A	PPB	24 Jul 2023 13:12:07	U	SMPL
O3680-02 - 1	TP-3	17	-0.0664	-	7470A	PPB	24 Jul 2023 13:14:23	U	SMPL
O3682-01 - 1	ELZ-23-00066-00070	33	-0.0626	-	7470A	PPB	24 Jul 2023 13:16:40	U	SMPL
O3682-02 - 1	ELZ-23-00061-00062-00069	46	-0.0596	-	7470A	PPB	24 Jul 2023 13:18:55	U	SMPL
CCV87 - 1	CCV87	22172	5.1421	-	7470A	PPB	24 Jul 2023 13:21:11	U	SMPL
CCB87 - 1	CCB87	-28	-0.0770	-	7470A	PPB	24 Jul 2023 13:23:26	U	SMPL
O3682-03 - 1	ELZ-23-00056-00057-00058	2	-0.0699	-	7470A	PPB	24 Jul 2023 13:25:41	U	SMPL
O3684-01 - 1	SF6-WASTE	18	-0.0662	-	7470A	PPB	24 Jul 2023 13:27:57	U	SMPL
O3684-01DUP - 1	SF6-WASTEDUP	0	-0.0704	-	7470A	PPB	24 Jul 2023 13:30:12	U	SMPL
O3684-01MS - 1	SF6-WASTEMS	16767	3.8715	-	7470A	PPB	24 Jul 2023 13:32:28	U	SMPL
O3684-01MSD - 1	SF6-WASTEMSD	17277	3.9914	-	7470A	PPB	24 Jul 2023 13:34:44	U	SMPL
PB154369BL - 1	PBW	-50	-0.0822	-	7470A	PPB	24 Jul 2023 13:37:00	U	SMPL
PB154369BS - 1	LCSW	17502	4.0442	-	7470A	PPB	24 Jul 2023 13:39:16	U	SMPL
O3722-10 - 1	TOTE-COMP	968	0.1572	-	7470A	PPB	24 Jul 2023 13:41:33	U	SMPL
O3722-12 - 1	60099	244	-0.0130	-	7470A	PPB	24 Jul 2023 13:43:51	U	SMPL
O3723-04 - 1	SUMP-1	-33	-0.0782	-	7470A	PPB	24 Jul 2023 13:46:07	U	SMPL
CCV88 - 1	CCV88	22331	5.1795	-	7470A	PPB	24 Jul 2023 13:48:23	U	SMPL
CCB88 - 1	CCB88	-67	-0.0861	-	7470A	PPB	24 Jul 2023 13:50:38	U	SMPL
O3723-04DUP - 1	SUMP-1DUP	21	-0.0655	-	7470A	PPB	24 Jul 2023 13:52:54	U	SMPL
O3723-04MS - 1	SUMP-1MS	17095	3.9486	-	7470A	PPB	24 Jul 2023 13:55:09	U	SMPL
O3723-04MSD - 1	SUMP-1MSD	17423	4.0257	-	7470A	PPB	24 Jul 2023 14:00:14	U	SMPL
PB154370BL - 1	PBW	-36	-0.0789	-	7470A	PPB	24 Jul 2023 14:02:30	U	SMPL
PB154370BS - 1	LCSW	14761	3.3999	-	7470A	PPB	24 Jul 2023 14:04:46	U	SMPL
O3682-04 - 1	ELZ-23-00059	50	-0.0586	-	7470A	PPB	24 Jul 2023 14:07:02	U	SMPL
O3682-04DUP - 1	ELZ-23-00059DUP	26	-0.0636	-	7470A	PPB	24 Jul 2023 14:09:19	U	SMPL
O3682-04MS - 1	ELZ-23-00059MS	16429	3.7920	-	7470A	PPB	24 Jul 2023 14:11:35	U	SMPL
O3682-04MSD - 1	ELZ-23-00059MSD	15953	3.6801	-	7470A	PPB	24 Jul 2023 14:13:52	U	SMPL
CCV89 - 1	CCV89	22246	5.1595	-	7470A	PPB	24 Jul 2023 14:18:52	U	SMPL
CCB89 - 1	CCB89	-100	-0.0939	-	7470A	PPB	24 Jul 2023 14:21:07	U	SMPL
PB154322TB - 1	PB154322TB	-19	-0.0749	-	7470A	PPB	24 Jul 2023 14:23:23	U	SMPL
PB154334TB - 1	PB154334TB	-4	-0.0713	-	7470A	PPB	24 Jul 2023 14:25:38	U	SMPL
O3684-01LX5 - 1		-31	-0.0777	-	7470A	PPB	24 Jul 2023 14:27:54	U	SMPL
O3684-01A - 1		16844	3.8896	-	7470A	PPB	24 Jul 2023 14:30:10	U	SMPL
O3723-04LX5 - 1		-30	-0.0775	-	7470A	PPB	24 Jul 2023 14:32:25	U	SMPL
O3723-04A - 1		18549	4.2904	-	7470A	PPB	24 Jul 2023 14:44:32	U	SMPL
O3682-04LX5 - 1		-36	-0.0789	-	7470A	PPB	24 Jul 2023 14:46:48	U	SMPL
O3682-04A - 1		16931	3.9100	-	7470A	PPB	24 Jul 2023 14:49:03	U	SMPL
CCV90 - 1	CCV90	21926	5.0843	-	7470A	PPB	24 Jul 2023 14:51:20	U	SMPL
CCB90 - 1	CCB90	-108	-0.0958	-	7470A	PPB	24 Jul 2023 14:53:38	U	SMPL