



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126571

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/24/2023	08:40
2	O3678-04		1	11	100.00	>212.0	>212.0	07/24/2023	09:20
3	O3678-04DUP		1	11	100.00	>212.0	>212.0	07/24/2023	09:50
4	O3718-06		1	12	100.00	>212.0	>212.0	07/24/2023	10:20
5	O3722-10		1	14	100.00	>212.0	>212.0	07/24/2023	10:50
6	O3722-12		1	12	100.00	>212.0	>212.0	07/24/2023	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126 571

WorkList Name : fp 7-24

WorkList ID : 172112

Department : Wet-Chemistry

Date : 07-24-2023 08:15:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3678-04 D	LWC-2	Water	Flash Point	Cool 4 deg C	REMI02		07/18/2023	1010B
O3718-06 A	DRIP-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/21/2023	1010B
O3722-10 B	TOTE-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/21/2023	1010B
O3722-12 D	60099	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/21/2023	1010B

Date/Time 07/24/2023 08:30
 Raw Sample Received by: RM1
 Raw Sample Relinquished by: RM1 (OC)

Date/Time 07/24/2023 12:00
 Raw Sample Received by: RM1 (OC)
 Raw Sample Relinquished by: RM1 (OC)