



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126572

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/24/2023	12:20
2	O3678-01		1	14	100.00	>212.0	>212.0	07/24/2023	12:50
3	O3678-03		1	11	100.00	>212.0	>212.0	07/24/2023	13:20
4	O3678-03DUP		1	11	100.00	>212.0	>212.0	07/24/2023	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126572

WorkList Name : FP-S-7-20

WorkList ID : 172063

Department : Wet-Chemistry

Date : 07-20-2023 15:11:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3678-01 C	LWC-1	Solid	Flash Point	Cool 4 deg C	REMI02	I11	07/18/2023	1010B
O3678-03 C	LWC-3	Solid	Flash Point	Cool 4 deg C	REMI02	I11	07/18/2023	1010B

Date/Time 07/24/2023 12:10
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: JH WCL

Date/Time 07/24/2023 14:30
 Raw Sample Received by: JH CWL
 Raw Sample Relinquished by: RM CWL