



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB126575

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/25/2023	08:05
2	CAL2	1	Water	NA	NA	20.5	7.01	07/25/2023	08:06
3	CAL3	1	Water	NA	NA	20.8	10.02	07/25/2023	08:09
4	ICV	1	Water	NA	NA	20.6	6.99	07/25/2023	08:10
5	CCV1	1	Water	NA	NA	20.6	2.01	07/25/2023	08:11
6	O3678-04	1	Water	NA	NA	20.9	8.79	07/25/2023	08:15
7	O3678-04DUP	1	Water	NA	NA	20.8	8.80	07/25/2023	08:16
8	CCV2	1	Water	NA	NA	20.7	12.02	07/25/2023	08:20

WORKLIST(Hardcopy Internal Chain)

W3 126575

WorkList Name : CORROSIONITY 03678

WorkList ID : 172147

Department : Wet-Chemistry

Date : 07-25-2023 07:54:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
03678-04	LWC-2	Water	Corrosivity	Cool 4 deg C	REMI02		07/18/2023	9040C

Date/Time 07/25/23 08:00
 Raw Sample Received by: SPC
 Raw Sample Relinquished by: CP SM

Date/Time 07/25/23 10:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: SPC