



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB126581

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	07/25/2023	08:15
2	CAL2	1	Water	NA	NA	20.6	7.01	07/25/2023	08:16
3	CAL3	1	Water	NA	NA	20.7	10.03	07/25/2023	08:17
4	ICV	1	Water	NA	NA	20.7	6.99	07/25/2023	08:20
5	CCV1	1	Water	NA	NA	20.7	2.01	07/25/2023	08:22
6	O3733-01	1	Solid	20.02	20	23.4	6.45	07/25/2023	08:25
7	O3733-01DUP	1	Solid	20.03	20	23.5	6.46	07/25/2023	08:26
8	CCV2	1	Water	NA	NA	20.7	12.02	07/25/2023	08:30

WORKLIST(Hardcopy Internal Chain)

U3126581

WorkList Name : PH 03733 S

WorkList ID : 172145

Department : Wet-Chemistry

Date : 07-25-2023 07:52:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3733-01	AU-6-072423	Solid	pH	Cool 4 deg C	PSEG05	J11	07/24/2023	9045D

Date/Time

07/25/23 08:00

Raw Sample Received by:

SA (CSC)

Raw Sample Relinquished by:

SA (CSC)

Date/Time

07/25/23

11:00

Raw Sample Received by:

CP SW

Raw Sample Relinquished by:

SA (CSC)