



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126582

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	07/25/2023	08:15
2	CAL2	1	Water	NA	NA	20.6	7.01	07/25/2023	08:16
3	CAL3	1	Water	NA	NA	20.7	10.03	07/25/2023	08:17
4	ICV	1	Water	NA	NA	20.7	6.99	07/25/2023	08:20
5	CCV1	1	Water	NA	NA	20.6	2.02	07/25/2023	08:35
6	O3735-02	1	Water	NA	NA	21.1	7.92	07/25/2023	08:40
7	O3735-02DUP	1	Water	NA	NA	21.2	7.93	07/25/2023	08:41
8	CCV2	1	Water	NA	NA	20.6	12.03	07/25/2023	08:45

WORKLIST(Hardcopy Internal Chain)

W126582

WorkList Name : PH W 03735

WorkList ID : 172146

Department : Wet-Chemistry

Date : 07-25-2023 07:53:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3735-02	1026	Water	pH	Cool 4 deg C	PSEG03	J11	07/24/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/25/23 08:00

JP (WOC)

JP

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/25/23

JP

JP

11:00

JP

JP