

LB126587


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

59456



A= 0.0000e+000

B= 1.7093e-004 *slope*

C= -1.0659e-001 *=intercept*

Rho= 0.9997505

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>g/b</i>
0.00	0.000	-0.001	-0.001	619	0.000	619					1
0.05	0.050					302	0				1
0.20	0.200	0.201	0.001	1801	0.0 %	1801					2
2.50	2.500	2.476	-0.024	15110	0.0 %	15110					2
5.00	5.000	5.119	0.119	30571	0.0 %	30571					2
7.50	7.500	7.349	-0.151	43616	0.0 %	43616					1
10.0	10.000	10.056	0.056	59456	0.0 %	59456					1

LB126587

INSTRUMENT IOD : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	619	-		0 SFAM01.1	PPB	7/25/2023 15:20	Std
	0.2 S01	1801	-		0.2 SFAM01.1	PPB	7/25/2023 15:22	Std
	2.5 S02	15110	-		2.5 SFAM01.1	PPB	7/25/2023 15:24	Std
	5 S03	30571	-		5 SFAM01.1	PPB	7/25/2023 15:26	Std
	7.5 S04	43616	-		7.5 SFAM01.1	PPB	7/25/2023 15:29	Std
	10 S05	59456	-		10 SFAM01.1	PPB	7/25/2023 15:31	Std
ICV083	ICV083	22972	3.8199 -		SFAM01.1	PPB	7/25/2023 15:34	SMPL
ICB083	ICB083	544	-0.0136 -		SFAM01.1	PPB	7/25/2023 15:36	SMPL
CCV091	CCV091	30437	5.0959 -		SFAM01.1	PPB	7/25/2023 15:38	SMPL
CCB091	CCB091	538	-0.0146 -		SFAM01.1	PPB	7/25/2023 15:41	SMPL
PB154422BL	PBW422	878	0.0435 -		SFAM01.1	PPB	7/25/2023 15:43	SMPL
O3693-01	EXEP3	797	0.0296 -		SFAM01.1	PPB	7/25/2023 15:45	SMPL
O3693-02	EXEP4	794	0.0291 -		SFAM01.1	PPB	7/25/2023 15:47	SMPL
O3693-03	EXEP5	840	0.037 -		SFAM01.1	PPB	7/25/2023 15:50	SMPL
O3693-04	EXEP5D	696	0.0124 -		SFAM01.1	PPB	7/25/2023 15:52	SMPL
O3693-05	EXEP5S	7080	1.1036 -		SFAM01.1	PPB	7/25/2023 15:54	SMPL
PB154423BL	PBW423	899	-0.0471 -		SFAM01.1	PPB	7/25/2023 15:57	SMPL
O3713-01	GCAW8	1015	0.0669 -		SFAM01.1	PPB	7/25/2023 15:59	SMPL
O3713-02	GCAW8D	1133	0.0871 -		SFAM01.1	PPB	7/25/2023 16:01	SMPL
O3713-03	GCAW8S	6724	1.0427 -		SFAM01.1	PPB	7/25/2023 16:03	SMPL
O3713-04	GCAW9	737	0.0194 -		SFAM01.1	PPB	7/25/2023 16:06	SMPL
O3713-05	GCAW0	806	0.0312 -		SFAM01.1	PPB	7/25/2023 16:08	SMPL
O3713-06	GCAW1	929	0.0522 -		SFAM01.1	PPB	7/25/2023 16:10	SMPL
PB154424BL	PBW424	795	0.0293 -		SFAM01.1	PPB	7/25/2023 16:15	SMPL
O3714-01	GCAW3	746	0.0209 -		SFAM01.1	PPB	7/25/2023 16:17	SMPL
O3714-02	GCAW3D	431	-0.0329 -		SFAM01.1	PPB	7/25/2023 16:19	SMPL
O3714-03	GCAW3S	6137	0.9424 -		SFAM01.1	PPB	7/25/2023 16:21	SMPL
O3714-04	GCAW4	615	-0.0015 -		SFAM01.1	PPB	7/25/2023 16:24	SMPL
O3714-05	GCAW5	624	0.0001 -		SFAM01.1	PPB	7/25/2023 16:26	SMPL
CCV092	CCV092	29418	4.9217 -		SFAM01.1	PPB	7/25/2023 16:28	SMPL
CCB092	CCB092	538	-0.0146 -		SFAM01.1	PPB	7/25/2023 16:31	SMPL
O3714-06	GCAW6	611	-0.0022 -		SFAM01.1	PPB	7/25/2023 16:33	SMPL
O3713-07	GCAW2	35614	5.9808 -		SFAM01.1	PPB	7/25/2023 16:37	SMPL
CCV093	CCV093	29514	4.9381 -		SFAM01.1	PPB	7/25/2023 16:40	SMPL
CCB093	CCB093	483	-0.024 -		SFAM01.1	PPB	7/25/2023 16:42	SMPL