

LB126589


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49845



A= 0.0000e+000

B= 1.9994e-004

C= -5.5552e-003

Rho= 0.9998217

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1
0.0	0.000	0.042	0.042	238	0.000	238	0				4
0.2	0.200	0.221	0.021	1134	0.0 %	1134					N
2.5	2.500	2.363	-0.137	11845	0.0 %	11845					-5
5.0	5.000	5.059	0.059	25329	0.0 %	25329					1
7.5	7.500	7.555	0.055	37816	0.0 %	37816					1
10.0	10.000	9.960	-0.040	49845	0.0 %	49845					0

LB126589 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jul 2023 08:45:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	238	-	0.0000	7470A	PPB	26 Jul 2023 09:33:22	S	Std
0.2 - 1	50.2	1134	-	0.2000	7470A	PPB	26 Jul 2023 09:35:38	S	Std
2.5 - 1	52.5	11845	-	2.5000	7470A	PPB	26 Jul 2023 09:37:56	S	Std
5.0 - 1	55	25329	-	5.0000	7470A	PPB	26 Jul 2023 09:40:13	S	Std
7.5 - 1	57.5	37816	-	7.5000	7470A	PPB	26 Jul 2023 09:45:00	S	Std
10.0 - 1	510	49845	-	10.0000	7470A	PPB	26 Jul 2023 09:49:50	S	Std
ICV84 - 1	ICV84	19573	3.9078	-	7470A	PPB	26 Jul 2023 09:52:37	U	SMPL
ICB84 - 1	ICB84	49	0.0042	-	7470A	PPB	26 Jul 2023 09:54:52	U	SMPL
CCV94 - 1	CCV94	22751	4.5432	-	7470A	PPB	26 Jul 2023 09:57:07	U	SMPL
CCB94 - 1	CCB94	98	0.0140	-	7470A	PPB	26 Jul 2023 09:59:22	U	SMPL
CRA - 1	CRA	1092	0.2128	-	7470A	PPB	26 Jul 2023 10:01:37	U	SMPL
HighStd - 1	HighStd	49811	9.9535	-	7470A	PPB	26 Jul 2023 10:03:52	U	SMPL
ChkStd - 1	ChkStd	35199	7.0320	-	7470A	PPB	26 Jul 2023 10:06:07	U	SMPL
PB154418BL - 1	PBW	50	0.0044	-	7470A	PPB	26 Jul 2023 10:08:23	U	SMPL
PB154418BS - 1	LCSW	17828	3.5589	-	7470A	PPB	26 Jul 2023 10:10:38	U	SMPL
O3717-02 - 1	TP-1	81	0.0106	-	7470A	PPB	26 Jul 2023 10:12:54	U	SMPL
O3717-05 - 1	TP-2	156	0.0256	-	7470A	PPB	26 Jul 2023 10:15:10	U	SMPL
O3738-02 - 1	80001	181	0.0306	-	7470A	PPB	26 Jul 2023 10:17:27	U	SMPL
O3738-05 - 1	80002	766	0.1476	-	7470A	PPB	26 Jul 2023 10:19:44	U	SMPL
O3742-01 - 1	WC-27	405	0.0754	-	7470A	PPB	26 Jul 2023 10:22:00	U	SMPL
CCV95 - 1	CCV95	25040	5.0008	-	7470A	PPB	26 Jul 2023 10:24:15	U	SMPL
CCB95 - 1	CCB95	137	0.0218	-	7470A	PPB	26 Jul 2023 10:26:30	U	SMPL
O3742-01DUP - 1	WC-27DUP	353	0.0650	-	7470A	PPB	26 Jul 2023 10:28:45	U	SMPL
O3742-01MS - 1	WC-27MS	20083	4.0098	-	7470A	PPB	26 Jul 2023 10:31:00	U	SMPL
O3742-01MSD - 1	WC-27MSD	18081	3.6095	-	7470A	PPB	26 Jul 2023 10:33:16	U	SMPL
PB154421BL - 1	PBW	75	0.0094	-	7470A	PPB	26 Jul 2023 10:35:31	U	SMPL
PB154421BS - 1	LCSW	19077	3.8086	-	7470A	PPB	26 Jul 2023 10:40:23	U	SMPL
O3678-04 - 1	LWC-2	1135	0.2214	-	7470A	PPB	26 Jul 2023 10:42:39	U	SMPL
O3735-02 - 1	1026	-4	-0.0064	-	7470A	PPB	26 Jul 2023 10:44:54	U	SMPL
O3735-02DUP - 1	1026DUP	102	0.0148	-	7470A	PPB	26 Jul 2023 10:47:13	U	SMPL
O3735-02MS - 1	1026MS	11425	2.2787	-	7470A	PPB	26 Jul 2023 10:52:20	U	SMPL
O3735-02MSD - 1	1026MSD	12139	2.4215	-	7470A	PPB	26 Jul 2023 10:54:35	U	SMPL
CCV96 - 1	CCV96	24673	4.9275	-	7470A	PPB	26 Jul 2023 10:56:50	U	SMPL
CCB96 - 1	CCB96	138	0.0220	-	7470A	PPB	26 Jul 2023 10:59:05	U	SMPL
PB154392TB - 1	PB154392TB	184	0.0312	-	7470A	PPB	26 Jul 2023 11:01:21	U	SMPL
O3742-01LX5 - 1		119	0.0182	-	7470A	PPB	26 Jul 2023 11:03:36	U	SMPL
O3742-01A - 1		19926	3.9784	-	7470A	PPB	26 Jul 2023 11:05:52	U	SMPL
O3735-02LX5 - 1		-17	-0.0090	-	7470A	PPB	26 Jul 2023 11:08:07	U	SMPL
O3735-02A - 1		15271	3.0477	-	7470A	PPB	26 Jul 2023 11:10:23	U	SMPL
CCV97 - 1	CCV97	24687	4.9303	-	7470A	PPB	26 Jul 2023 11:12:39	U	SMPL
CCB97 - 1	CCB97	116	0.0176	-	7470A	PPB	26 Jul 2023 11:14:56	U	SMPL