

LB126594


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

6363E



A= 0.0000e+000

B= 1.5605e-004

C= -4.4107e-002

Rho= 0.9997193

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.036	-0.036	55	0.000	55					12
0.2	0.200	0.224	0.024	1718	0.0 %	1718					0
2.5	2.500	2.488	-0.012	16224	0.0 %	16224					-1
5.0	5.000	4.966	-0.034	32104	0.0 %	32104					2
7.5	7.500	7.672	0.172	49449	0.0 %	49449					-1
10.0	10.000	9.886	-0.114	63635	0.0 %	63635					

LB126594 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jul 2023 11:29:17

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	55	-	0.0000	7471B	PPB	26 Jul 2023 11:31:06	S	Std
0.2 - 1	50.2	1718	-	0.2000	7471B	PPB	26 Jul 2023 11:33:23	S	Std
2.5 - 1	52.5	16224	-	2.5000	7471B	PPB	26 Jul 2023 11:35:39	S	Std
5.0 - 1	55	32104	-	5.0000	7471B	PPB	26 Jul 2023 11:37:56	S	Std
7.5 - 1	57.5	49449	-	7.5000	7471B	PPB	26 Jul 2023 11:40:12	S	Std
10.0 - 1	510	63635	-	10.0000	7471B	PPB	26 Jul 2023 11:42:28	S	Std
ICV85 - 1	ICV85	25147	3.8800	-	7471B	PPB	26 Jul 2023 11:47:49	U	SMPL
ICB85 - 1	ICB85	-2	-0.0444	-	7471B	PPB	26 Jul 2023 11:50:04	U	SMPL
CCV01 - 1	CCV01	32580	5.0399	-	7471B	PPB	26 Jul 2023 11:52:19	U	SMPL
CCB01 - 1	CCB01	-21	-0.0474	-	7471B	PPB	26 Jul 2023 11:54:34	U	SMPL
CRA - 1	CRA	1698	0.2209	-	7471B	PPB	26 Jul 2023 11:56:50	U	SMPL
HighStd - 1	HighStd	63949	9.9350	-	7471B	PPB	26 Jul 2023 11:59:05	U	SMPL
ChkStd - 1	ChkStd	47211	7.3231	-	7471B	PPB	26 Jul 2023 12:01:19	U	SMPL
PB154416BL - 1	PBS	-36	-0.0497	-	7471B	PPB	26 Jul 2023 12:03:37	U	SMPL
PB154416BS - 1	LCSS	24229	3.7368	-	7471B	PPB	26 Jul 2023 12:05:54	U	SMPL
O3686-09 - 1	SB-29-0-2-20230721	46290	7.1793	-	7471B	PPB	26 Jul 2023 12:13:07	U	SMPL
O3686-09DUP - 1	SB-29-0-2-20230721DUP	45872	7.1141	-	7471B	PPB	26 Jul 2023 12:15:23	U	SMPL
O3686-09MS - 1	SB-29-0-2-20230721MS	72095	11.2062	-	7471B	PPB	26 Jul 2023 12:17:41	U	SMPL
O3686-09MSD - 1	SB-29-0-2-20230721MSD	72752	11.3087	-	7471B	PPB	26 Jul 2023 12:19:58	U	SMPL
O3686-11 - 1	SB-23-0-2-20230721	63965	9.9375	-	7471B	PPB	26 Jul 2023 12:22:14	U	SMPL
CCV02 - 1	CCV02	31974	4.9454	-	7471B	PPB	26 Jul 2023 12:24:31	U	SMPL
CCB02 - 1	CCB02	20	-0.0410	-	7471B	PPB	26 Jul 2023 12:26:48	U	SMPL
O3686-13 - 1	SB-22-0-2-20230721	26198	0.4040	-	7471B	PPB	26 Jul 2023 12:29:03	U	SMPL
O3686-14 - 1	SB-22-4-6-20230721	2721	0.3805	-	7471B	PPB	26 Jul 2023 12:31:18	U	SMPL
O3686-16 - 1	SB-28-0-2-20230721	31593	4.8859	-	7471B	PPB	26 Jul 2023 12:33:34	U	SMPL
O3724-02 - 1	185465	4316	0.6294	-	7471B	PPB	26 Jul 2023 12:35:50	U	SMPL
O3724-04 - 1	363	5309	0.7844	-	7471B	PPB	26 Jul 2023 12:38:05	U	SMPL
O3731-01 - 1	WC-1	92549	14.3979	-	7471B	PPB	26 Jul 2023 12:40:21	U	SMPL
O3733-01 - 1	AU-6-072423	9158	1.3850	-	7471B	PPB	26 Jul 2023 12:42:37	U	SMPL
O3735-01 - 1	1029	675	0.0612	-	7471B	PPB	26 Jul 2023 12:44:56	U	SMPL
O3736-01 - 1	TR-5-072423	1534	0.1953	-	7471B	PPB	26 Jul 2023 12:47:12	U	SMPL
O3738-01 - 1	80001	643	0.0562	-	7471B	PPB	26 Jul 2023 12:49:28	U	SMPL
CCV03 - 1	CCV03	32985	5.1031	-	7471B	PPB	26 Jul 2023 12:51:43	U	SMPL
CCB03 - 1	CCB03	32	-0.0391	-	7471B	PPB	26 Jul 2023 12:53:58	U	SMPL
O3738-04 - 1	80002	287	0.0007	-	7471B	PPB	26 Jul 2023 12:56:14	U	SMPL
O3742-02 - 1	B-P32A	5477	0.8106	-	7471B	PPB	26 Jul 2023 12:58:29	U	SMPL
O3742-03 - 1	B-P32B	3016	0.4265	-	7471B	PPB	26 Jul 2023 13:00:45	U	SMPL
O3742-04 - 1	B-P33A	4706	0.6903	-	7471B	PPB	26 Jul 2023 13:03:00	U	SMPL
O3742-05 - 1	B-P33B	3012	0.4259	-	7471B	PPB	26 Jul 2023 13:05:16	U	SMPL
O3742-06 - 1	DUP-02	2589	0.3599	-	7471B	PPB	26 Jul 2023 13:07:32	U	SMPL
O3742-07 - 1	DUP-03	2306	0.3157	-	7471B	PPB	26 Jul 2023 13:09:48	U	SMPL
O3742-08 - 1	DUP-04	2330	0.3195	-	7471B	PPB	26 Jul 2023 13:12:05	U	SMPL
O3686-09LX5 - 1		10348	1.5707	-	7471B	PPB	26 Jul 2023 13:14:22	U	SMPL
O3686-09A - 1		72827	11.3204	-	7471B	PPB	26 Jul 2023 13:16:37	U	SMPL
CCV04 - 1	CCV04	32961	5.0994	-	7471B	PPB	26 Jul 2023 13:18:53	U	SMPL
CCB04 - 1	CCB04	5	-0.0433	-	7471B	PPB	26 Jul 2023 13:21:10	U	SMPL
O3686-09DLX2 - 1		22188	3.4183	-	7471B	PPB	26 Jul 2023 13:23:25	U	SMPL
O3686-09DUPDLX2 - 1		24030	3.7057	-	7471B	PPB	26 Jul 2023 13:25:40	U	SMPL
O3686-09MSDLX2 - 1		37686	5.8367	-	7471B	PPB	26 Jul 2023 13:27:56	U	SMPL
O3686-09MSDDLX2 - 1		36824	5.7022	-	7471B	PPB	26 Jul 2023 13:30:11	U	SMPL
O3686-09LDLX10 - 1		5095	0.7510	-	7471B	PPB	26 Jul 2023 13:32:27	U	SMPL
O3686-09ADLX2 - 1		37220	5.7640	-	7471B	PPB	26 Jul 2023 13:34:43	U	SMPL
O3731-01DLX5 - 1		19837	3.0514	-	7471B	PPB	26 Jul 2023 13:36:59	U	SMPL
CCV05 - 1	CCV05	33471	5.1790	-	7471B	PPB	26 Jul 2023 13:39:16	U	SMPL
CCB05 - 1	CCB05	27	-0.0399	-	7471B	PPB	26 Jul 2023 13:41:33	U	SMPL