



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB126597

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/26/2023	10:30
2	CAL2	1	Water	NA	NA	20.7	7.00	07/26/2023	10:31
3	CAL3	1	Water	NA	NA	20.7	10.02	07/26/2023	10:33
4	ICV	1	Water	NA	NA	20.6	6.99	07/26/2023	10:35
5	CCV1	1	Water	NA	NA	20.7	2.01	07/26/2023	10:40
6	O3746-01	1	Water	NA	NA	23.3	10.46	07/26/2023	10:44
7	O3752-01	1	Water	NA	NA	23.4	4.86	07/26/2023	10:45
8	O3752-02	1	Water	NA	NA	23.7	4.37	07/26/2023	10:46
9	O3752-03	1	Water	NA	NA	23.4	5.29	07/26/2023	10:47
10	O3766-01	1	Water	NA	NA	23.5	4.94	07/26/2023	10:50
11	O3766-02	1	Water	NA	NA	23.6	4.90	07/26/2023	10:52
12	O3767-01	1	Water	NA	NA	20.2	10.05	07/26/2023	11:00
13	O3771-01	1	Water	NA	NA	20.3	4.85	07/26/2023	11:10
14	O3771-01DUP	1	Water	NA	NA	20.4	4.86	07/26/2023	11:11
15	CCV2	1	Water	NA	NA	20.7	12.02	07/26/2023	11:15

WORKLIST(Hardcopy Internal Chain)

WB12659f

WorkList Name : ph o3770

WorkList ID : 172197

Department : Wet-Chemistry

Date : 07-26-2023 10:11:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3746-01	A A455	Water	pH	Cool 4 deg C	PSEG03	J11	07/25/2023	9040C
O3752-01	A CELL-A	Water	pH	Cool 4 deg C	STER10	D11	07/21/2023	9040C
O3752-02	A CELL-B/C	Water	pH	Cool 4 deg C	STER10	D11	07/21/2023	9040C
O3752-03	A CELL-D	Water	pH	Cool 4 deg C	STER10	D11	07/21/2023	9040C
O3766-01	A MAIN-CELL	Water	pH	Cool 4 deg C	STER16	J11	07/24/2023	9040C
O3766-02	A EX-CELL	Water	pH	Cool 4 deg C	STER16	J11	07/24/2023	9040C
O3767-01	E MMC-COMP	Water	pH	Cool 4 deg C	PSEG03		07/25/2023	9040C
O3771-01	E PATERSON-GSMP-COMP	Water	pH	Cool 4 deg C	PSEG03		07/25/2023	9040C

Date/Time 07/26/23 10:20

Raw Sample Received by: SD (WDC)

Raw Sample Relinquished by: SD (WDC)

Date/Time 07/26/23

Raw Sample Received by: SD (WDC)

Raw Sample Relinquished by: SD (WDC)