

LB126605


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

65156



A= 0.0000e+000

B= 1.5232e-004

C= -5.5387e-002

Rho= 0.9997278

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	-0.043	-0.043	81	0.000	81					6
0.2	0.200	0.212	0.012	1753	0.0 %	1753					-2
2.5	2.500	2.457	-0.043	16495	0.0 %	16495					1
5.0	5.000	5.068	0.068	33635	0.0 %	33635					2
7.5	7.500	7.637	0.137	50505	0.0 %	50505					-1
10.0	10.000	9.869	-0.131	65156	0.0 %	65156					

LB126605 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jul 2023 15:43:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	81	-	0.0000	7471B	PPB	26 Jul 2023 15:45:29	S	Std
0.2 - 1	50.2	1753	-	0.2000	7471B	PPB	26 Jul 2023 15:47:46	S	Std
2.5 - 1	52.5	16495	-	2.5000	7471B	PPB	26 Jul 2023 15:50:03	S	Std
5.0 - 1	55	33635	-	5.0000	7471B	PPB	26 Jul 2023 15:52:19	S	Std
7.5 - 1	57.5	50505	-	7.5000	7471B	PPB	26 Jul 2023 15:54:35	S	Std
10.0 - 1	510	65156	-	10.0000	7471B	PPB	26 Jul 2023 15:56:51	S	Std
ICV86 - 1	ICV86	25901	3.8898	-7471B	PPB	26 Jul 2023 16:02:22	U	SMPL	
ICB86 - 1	ICB86	55	-0.0470	-7471B	PPB	26 Jul 2023 16:04:37	U	SMPL	
CCV06 - 1	CCV06	33172	4.9973	-7471B	PPB	26 Jul 2023 16:06:52	U	SMPL	
CCB06 - 1	CCB06	35	-0.0501	-7471B	PPB	26 Jul 2023 16:09:07	U	SMPL	
CRA - 1	CRA	1722	0.2069	-7471B	PPB	26 Jul 2023 16:11:22	U	SMPL	
HighStd - 1	HighStd	65223	9.8793	-7471B	PPB	26 Jul 2023 16:13:37	U	SMPL	
ChkStd - 1	ChkStd	47808	7.2266	-7471B	PPB	26 Jul 2023 16:15:52	U	SMPL	
PB154417BL - 1	PBS	-30	-0.0600	-7471B	PPB	26 Jul 2023 16:18:09	U	SMPL	
PB154417BS - 1	LCSS	24376	3.6575	-7471B	PPB	26 Jul 2023 16:20:26	U	SMPL	
O3740-01 - 1	SU-04-072423	2363	0.3045	-7471B	PPB	26 Jul 2023 16:22:42	U	SMPL	
O3741-01 - 1	B-P5A	4392	0.6136	-7471B	PPB	26 Jul 2023 16:24:59	U	SMPL	
O3741-02 - 1	B-P12A	1350	0.1502	-7471B	PPB	26 Jul 2023 16:27:15	U	SMPL	
O3741-03 - 1	B-P12B	1245	0.1342	-7471B	PPB	26 Jul 2023 16:29:32	U	SMPL	
O3741-04 - 1	B-P24A	2241	0.2860	-7471B	PPB	26 Jul 2023 16:31:46	U	SMPL	
CCV07 - 1	CCV07	33377	5.0285	-7471B	PPB	26 Jul 2023 16:34:01	U	SMPL	
CCB07 - 1	CCB07	70	-0.0447	-7471B	PPB	26 Jul 2023 16:36:17	U	SMPL	
O3741-05 - 1	B-P24B	2	-0.0551	-7471B	PPB	26 Jul 2023 16:38:32	U	SMPL	
O3741-06 - 1	B-P25A	17180	2.5614	-7471B	PPB	26 Jul 2023 16:40:48	U	SMPL	
O3741-07 - 1	B-P25B	849	0.0739	-7471B	PPB	26 Jul 2023 16:43:03	U	SMPL	
O3741-08 - 1	B-P26A	834	0.0716	-7471B	PPB	26 Jul 2023 16:45:18	U	SMPL	
O3741-09 - 1	B-P26B	184	-0.0274	-7471B	PPB	26 Jul 2023 16:47:35	U	SMPL	
O3741-10 - 1	B-P27A	725	0.0550	-7471B	PPB	26 Jul 2023 16:49:53	U	SMPL	
O3741-11 - 1	B-P27B	-40	-0.0615	-7471B	PPB	26 Jul 2023 16:52:11	U	SMPL	
O3741-12 - 1	B-P6A	928	0.0860	-7471B	PPB	26 Jul 2023 16:54:28	U	SMPL	
O3741-13 - 1	B-P6B	1837	0.2244	-7471B	PPB	26 Jul 2023 16:56:44	U	SMPL	
O3741-14 - 1	B-P13A	4705	0.6613	-7471B	PPB	26 Jul 2023 16:59:01	U	SMPL	
CCV08 - 1	CCV08	33668	5.0729	-7471B	PPB	26 Jul 2023 17:16:09	U	SMPL	
CCB08 - 1	CCB08	64	-0.0456	-7471B	PPB	26 Jul 2023 17:18:24	U	SMPL	
O3741-15 - 1	B-P13B	2079	0.2613	-7471B	PPB	26 Jul 2023 17:20:39	U	SMPL	
O3741-16 - 1	B-P28A	1636	0.1938	-7471B	PPB	26 Jul 2023 17:22:55	U	SMPL	
O3741-17 - 1	B-P28B	731	0.0560	-7471B	PPB	26 Jul 2023 17:25:10	U	SMPL	
O3741-18 - 1	B-P29A	2486	0.3233	-7471B	PPB	26 Jul 2023 17:27:25	U	SMPL	
O3741-19 - 1	B-P29B	808	0.0677	-7471B	PPB	26 Jul 2023 17:29:41	U	SMPL	
O3741-19DUP - 1	B-P29BDUP	632	0.0409	-7471B	PPB	26 Jul 2023 17:31:58	U	SMPL	
O3741-19MS - 1	B-P29BMS	28535	4.2910	-7471B	PPB	26 Jul 2023 17:34:14	U	SMPL	
O3741-19MSD - 1	B-P29BMDS	27698	4.1635	-7471B	PPB	26 Jul 2023 17:36:31	U	SMPL	
O3741-19LX5 - 1		152	-0.0322	-7471B	PPB	26 Jul 2023 17:38:48	U	SMPL	
O3741-19A - 1		27815	4.1813	-7471B	PPB	26 Jul 2023 17:41:04	U	SMPL	
CCV09 - 1	CCV09	34052	5.1313	-7471B	PPB	26 Jul 2023 17:43:20	U	SMPL	
CCB09 - 1	CCB09	54	-0.0472	-7471B	PPB	26 Jul 2023 17:45:36	U	SMPL	