



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126606

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	07/26/2023	11:50
2	O3735-02		1	11	100.00	>212.0	>212.0	07/26/2023	12:20
3	O3735-02DUP		1	13	100.00	>212.0	>212.0	07/26/2023	12:50
4	O3746-01		1	14	100.00	>212.0	>212.0	07/26/2023	13:20
5	O3767-01		1	12	100.00	>212.0	>212.0	07/26/2023	13:50
6	O3771-01		1	14	100.00	>212.0	>212.0	07/26/2023	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126606

WorkList Name : fp 7-25 WorkList ID : 172152 Department : Wet-Chemistry Date : 07-25-2023 08:02:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3735-02C	1026	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/24/2023	1010B
O3746-01D	A455	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/25/2023	1010B
O3767-01I	MMC-COMP	Water	Flash Point	Cool 4 deg C	PSEG03		07/25/2023	1010B
O3771-01I	PATERSON-GSMP-COMP	Water	Flash Point	Cool 4 deg C	PSEG03		07/25/2023	1010B

Date/Time 07/26/2023 11:45
 Raw Sample Received by: [Signature] CWC
 Raw Sample Relinquished by: [Signature] CWC

Date/Time 07/26/2023 15:00
 Raw Sample Received by: [Signature] CWC
 Raw Sample Relinquished by: [Signature] CWC