



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126609

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/26/2023	15:10
2	CAL2	1	Water	NA	NA	20.7	7.00	07/26/2023	15:11
3	CAL3	1	Water	NA	NA	20.6	10.02	07/26/2023	15:12
4	ICV	1	Water	NA	NA	20.6	7.00	07/26/2023	15:15
5	CCV1	1	Water	NA	NA	20.7	2.01	07/26/2023	15:20
6	O3756-01	1	Solid	20.02	20	22.5	9.23	07/26/2023	15:22
7	O3756-01DUP	1	Solid	20.03	20	22.6	9.24	07/26/2023	15:23
8	O3756-04	1	Solid	20.04	20	22.2	9.30	07/26/2023	15:25
9	O3756-07	1	Solid	20.01	20	22.1	9.17	07/26/2023	15:30
10	O3756-10	1	Solid	20.04	20	22.3	8.94	07/26/2023	15:33
11	O3756-13	1	Solid	20.02	20	22.2	9.09	07/26/2023	15:35
12	O3756-16	1	Solid	20.03	20	22.2	9.29	07/26/2023	15:40
13	O3756-20	1	Solid	20.01	20	22.1	8.99	07/26/2023	15:45
14	O3756-24	1	Solid	20.02	20	22.2	8.55	07/26/2023	15:50
15	O3768-02	1	Solid	20.01	20	20.1	5.71	07/26/2023	16:00
16	CCV2	1	Water	NA	NA	20.7	12.02	07/26/2023	16:02
17	O3770-02	1	Solid	20.01	20	20.2	6.05	07/26/2023	16:10
18	O3770-05	1	Solid	20.03	20	20.1	6.52	07/26/2023	16:15
19	O3770-08	1	Solid	20.02	20	20.1	7.19	07/26/2023	16:20
20	O3770-11	1	Solid	20.03	20	20.1	6.86	07/26/2023	16:25

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O3784-03	1	Solid	20.01	20	22.7	5.30	07/26/2023	16:30
22	CCV3	1	Water	NA	NA	20.7	2.02	07/26/2023	16:38

WORKLIST(Hardcopy Internal Chain)

17126609

WorkList Name : CORROSIONITY O3770

WorkList ID : 172200

Department : Wet-Chemistry

Date : 07-26-2023 10:13:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3756-01	RPTP-11-CS-11	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3756-04	RPTP-10-CS-10	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3756-07	RRTP-9-CS-9	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3756-10	RRTP-14-CS-14	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3756-13	RRTP-13-CS-13	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3756-16	RRTP-12-CS-12	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3756-20	307/308-TP-6-8-CS3	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3756-24	415/417-TP-1-4-CS-1	Solid	Corrosivity	Cool 4 deg C	PREF02	J11	07/21/2023	9045D
O3768-02	FL-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	J11	07/25/2023	9045D
O3770-02	DEL-1-RP	Solid	Corrosivity	Cool 4 deg C	PSEG03	J11	07/25/2023	9045D
O3770-05	DEL-2-RP	Solid	Corrosivity	Cool 4 deg C	PSEG03	J11	07/25/2023	9045D
O3770-08	HOWE-1-LP	Solid	Corrosivity	Cool 4 deg C	PSEG03	J11	07/25/2023	9045D
O3770-11	HOWE-2-LP	Solid	Corrosivity	Cool 4 deg C	PSEG03	J11	07/25/2023	9045D
O3784-03	OD-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	J11	07/26/2023	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/26/23 15:00

CPG

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/26/23

CPG