

LB12640


7470A

INSTRUMENT 1A: CV1

Linear

 μ Abs.:

46115



A= 0.0000e+000

B= 2.1657e-004

C= -4.8008e-002

Rho= 0.9999324

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	b.o.d
0.0	0.000	-0.012	-0.012	165	0.000	165	0				
0.2	0.200	0.199	-0.001	1141	0.0 %	1141					
2.5	2.500	2.465	-0.035	11605	0.0 %	11605					
5.0	5.000	5.052	0.052	23551	0.0 %	23551					
7.5	7.500	7.557	0.057	35115	0.0 %	35115					
10.0	10.000	9.939	-0.061	46115	0.0 %	46115					

LB126640 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Jul 2023 11:12:37

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	165	-	0.0000	7470A	PPB	28 Jul 2023 11:23:27	S	Std
0.2 - 1	50.2	1141	-	0.2000	7470A	PPB	28 Jul 2023 11:25:44	S	Std
2.5 - 1	52.5	11605	-	2.5000	7470A	PPB	28 Jul 2023 11:28:01	S	Std
5.0 - 1	55	23551	-	5.0000	7470A	PPB	28 Jul 2023 11:30:18	S	Std
7.5 - 1	57.5	35115	-	7.5000	7470A	PPB	28 Jul 2023 11:35:10	S	Std
10.0 - 1	510	46115	-	10.0000	7470A	PPB	28 Jul 2023 11:39:55	S	Std
ICV89 - 1	ICV89	18291	3.9132	-	7470A	PPB	28 Jul 2023 11:42:50	U	SMPL
ICB89 - 1	ICB89	36	-0.0402	-	7470A	PPB	28 Jul 2023 11:45:04	U	SMPL
CCV20 - 1	CCV20	23283	4.9943	-	7470A	PPB	28 Jul 2023 11:47:19	U	SMPL
CCB20 - 1	CCB20	32	-0.0411	-	7470A	PPB	28 Jul 2023 11:49:34	U	SMPL
CRA - 1	CRA	1098	0.1898	-	7470A	PPB	28 Jul 2023 11:51:49	U	SMPL
HighStd - 1	HighStd	46132	9.9426	-	7470A	PPB	28 Jul 2023 11:54:04	U	SMPL
ChkStd - 1	ChkStd	32657	7.0244	-	7470A	PPB	28 Jul 2023 11:56:20	U	SMPL
PB154487BL - 1	PBW	-93	-0.0681	-	7470A	PPB	28 Jul 2023 11:58:36	U	SMPL
PB154487BS - 1	LCSW	16203	3.4610	-	7470A	PPB	28 Jul 2023 12:00:52	U	SMPL
O3768-02 - 1	FL-1	-52	-0.0593	-	7470A	PPB	28 Jul 2023 12:03:08	U	SMPL
O3770-02 - 1	DEL-1-RP	30	-0.0415	-	7470A	PPB	28 Jul 2023 12:05:24	U	SMPL
O3770-05 - 1	DEL-2-RP	86	-0.0294	-	7470A	PPB	28 Jul 2023 12:07:41	U	SMPL
O3770-08 - 1	HOWE-1-LP	43	-0.0387	-	7470A	PPB	28 Jul 2023 12:09:58	U	SMPL
O3770-11 - 1	HOWE-2-LP	33	-0.0409	-	7470A	PPB	28 Jul 2023 12:12:14	U	SMPL
CCV21 - 1	CCV21	23432	5.0266	-	7470A	PPB	28 Jul 2023 12:14:29	U	SMPL
CCB21 - 1	CCB21	54	-0.0363	-	7470A	PPB	28 Jul 2023 12:16:45	U	SMPL
O3770-11DUP - 1	HOWE-2-LPDUP	26	-0.0424	-	7470A	PPB	28 Jul 2023 12:19:00	U	SMPL
O3770-11MS - 1	HOWE-2-LPMS	16727	3.5745	-	7470A	PPB	28 Jul 2023 12:21:15	U	SMPL
O3770-11MSD - 1	HOWE-2-LPMSD	15830	3.3802	-	7470A	PPB	28 Jul 2023 12:23:30	U	SMPL
PB154442TB - 1	PB154442TB	-42	-0.0571	-	7470A	PPB	28 Jul 2023 12:25:46	U	SMPL
O3770-11LX5 - 1		-7	-0.0495	-	7470A	PPB	28 Jul 2023 12:28:02	U	SMPL
O3770-11A - 1		17381	3.7161	-	7470A	PPB	28 Jul 2023 12:30:18	U	SMPL
CCV22 - 1	CCV22	23382	5.0157	-	7470A	PPB	28 Jul 2023 12:32:34	U	SMPL
CCB22 - 1	CCB22	38	-0.0398	-	7470A	PPB	28 Jul 2023 12:34:50	U	SMPL