



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126645

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	07/28/2023	12:40
2	O3806-01		1	12	100.00	>212.0	>212.0	07/28/2023	13:20
3	O3806-02		1	14	100.00	>212.0	>212.0	07/28/2023	13:50
4	O3806-03		1	11	100.00	>212.0	>212.0	07/28/2023	14:20
5	O3808-01		1	13	100.00	>212.0	>212.0	07/28/2023	14:50
6	O3808-09		1	14	100.00	>212.0	>212.0	07/28/2023	15:20
7	O3819-02		1	14	100.00	>212.0	>212.0	07/28/2023	15:50
8	O3819-02DUP		1	12	100.00	>212.0	>212.0	07/28/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6645

WorkList Name : FP-W-7-27

WorkList ID : 172265

Department : Wet-Chemistry

Date : 07-27-2023 12:21:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3806-01 E	ORA-1694	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/27/2023	1010B
O3806-02 C	ORA-1695	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/27/2023	1010B
O3806-03 C	ORA-1693	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/27/2023	1010B
O3808-01 F	1709	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/27/2023	1010B
O3808-09 C	1710-1715	Water	Flash Point	Cool 4 deg C	PSEG03	J11	07/27/2023	1010B
O3819-02 C	370	Water	Flash Point	Cool 4 deg C	PSEG03	L11	07/27/2023	1010B

Date/Time 07/28/2023 12:30
 Raw Sample Received by: RM awc
 Raw Sample Relinquished by: JG (awc)

Date/Time 07/28/2023 16:30
 Raw Sample Received by: JG (awc)
 Raw Sample Relinquished by: RM (awc)