

LB126646


SFAM01.1

INSTRUMENT ID: CVM

Linear

 μ Abs.:

54073



A= 0.0000e+000

B= 1.8504e-004 slope

C= -5.0451e-002 y-intercept

Rho= 0.9997782

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/6/13
0.00	0.000	-0.012	-0.012	207	0.000	207					14
0.05	0.050	0.043	-0.007	504	0.0 %	504	0				12
0.20	0.200	0.197	-0.003	1338	0.0 %	1338					13
2.50	2.500	2.428	-0.072	13393	0.0 %	13393					4
5.00	5.000	5.183	0.183	28283	0.0 %	28283					1
7.50	7.500	7.457	-0.043	40571	0.0 %	40571					0
10.0	10.000	9.955	-0.045	54073	0.0 %	54073					

LB126646

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	207	-		0 SFAM01.1	PPB	7/28/2023 13:35	Std
	0.05 S01	504	-		0.05 SFAM01.1	PPB	7/28/2023 13:37	Std
	0.2 S02	1338	-		0.2 SFAM01.1	PPB	7/28/2023 13:39	Std
	2.5 S03	13393	-		2.5 SFAM01.1	PPB	7/28/2023 13:42	Std
	5 S04	28283	-		5 SFAM01.1	PPB	7/28/2023 13:44	Std
	7.5 S05	40571	-		7.5 SFAM01.1	PPB	7/28/2023 13:46	Std
	10 S06	54073	-		10 SFAM01.1	PPB	7/28/2023 13:51	Std
ICV090	ICV090	22544	4.121 -		SFAM01.1	PPB	7/28/2023 13:54	SMPL
ICB090	ICB090	229	-0.0081 -		SFAM01.1	PPB	7/28/2023 13:59	SMPL
CCV023	CCV023	28350	5.1953 -		SFAM01.1	PPB	7/28/2023 14:01	SMPL
CCB023	CCB023	156	-0.0216 -		SFAM01.1	PPB	7/28/2023 14:03	SMPL
PB154518BL	PBW518	212	-0.0112 -		SFAM01.1	PPB	7/28/2023 14:08	SMPL
O3693-01	EXEP3	757	0.0896 -		SFAM01.1	PPB	7/28/2023 14:10	SMPL
O3693-02	EXEP4	766	0.0913 -		SFAM01.1	PPB	7/28/2023 14:13	SMPL
O3693-03	EXEP5	539	0.0493 -		SFAM01.1	PPB	7/28/2023 14:18	SMPL
O3693-04	EXEP5D	526	0.0469 -		SFAM01.1	PPB	7/28/2023 14:20	SMPL
O3693-05	EXEP5S	6848	1.2167 -		SFAM01.1	PPB	7/28/2023 14:22	SMPL
CCV024	CCV024	28053	5.1403 -		SFAM01.1	PPB	7/28/2023 14:24	SMPL
CCB024	CCB024	202	-0.0131 -		SFAM01.1	PPB	7/28/2023 14:29	SMPL