


LB126658

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0879e-004

C= 1.7304e-002

Rho= 0.9999282

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/01
0.0	0.000	0.040	0.040	107	0.000	107	0				12
0.2	0.200	0.223	0.023	985	0.0 %	985					13
2.5	2.500	2.431	-0.069	11561	0.0 %	11561					11
5.0	5.000	4.955	-0.045	23648	0.0 %	23648					11
7.5	7.500	7.553	0.053	36094	0.0 %	36094					0
10.0	10.000	9.998	-0.002	47804	0.0 %	47804					

LB126658 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Jul 2023 10:32:06

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	107	-	0.0000	7470A	PPB	31 Jul 2023 10:47:29	S	Std
0.2 - 1	50.2	985	-	0.2000	7470A	PPB	31 Jul 2023 10:49:45	S	Std
2.5 - 1	52.5	11561	-	2.5000	7470A	PPB	31 Jul 2023 10:52:02	S	Std
5.0 - 1	55	23648	-	5.0000	7470A	PPB	31 Jul 2023 10:54:19	S	Std
7.5 - 1	57.5	36094	-	7.5000	7470A	PPB	31 Jul 2023 10:56:35	S	Std
10.0 - 1	57.1	47804	-	10.0000	7470A	PPB	31 Jul 2023 10:58:51	S	Std
ICV92 - 1	ICV92	19790	4.1492	-	7470A	PPB	31 Jul 2023 11:01:54	U	SMPL
ICB92 - 1	ICB92	-29	0.0112	-	7470A	PPB	31 Jul 2023 11:04:08	U	SMPL
CCV29 - 1	CCV29	23212	4.8637	-	7470A	PPB	31 Jul 2023 11:06:23	U	SMPL
CCB29 - 1	CCB29	-34	0.0102	-	7470A	PPB	31 Jul 2023 11:08:38	U	SMPL
CRA - 1	CRA	925	0.2104	-	7470A	PPB	31 Jul 2023 11:10:53	U	SMPL
HighStd - 1	HighStd	47804	9.9982	-	7470A	PPB	31 Jul 2023 11:13:11	U	SMPL
ChkStd - 1	ChkStd	34146	7.1486	-	7470A	PPB	31 Jul 2023 11:15:26	U	SMPL
PB154514BL - 1	PBW	-65	0.0037	-	7470A	PPB	31 Jul 2023 11:17:42	U	SMPL
PB154514BS - 1	LCSW	18685	3.9185	-	7470A	PPB	31 Jul 2023 11:19:57	U	SMPL
O3795-02 - 1	TOPSOIL-OU3-DENALI-TEMP	103	0.0388	-	7470A	PPB	31 Jul 2023 11:22:13	U	SMPL
O3796-01 - 1	TOPSOIL-OU3-DENALI-TEMP	65	0.0309	-	7470A	PPB	31 Jul 2023 11:24:30	U	SMPL
O3796-01DUP - 1	TOPSOIL-OU3-DENALI-TEMPDUP	46	0.0269	-	7470A	PPB	31 Jul 2023 11:26:47	U	SMPL
O3796-01MS - 1	TOPSOIL-OU3-DENALI-TEMPMS	16537	3.4700	-	7470A	PPB	31 Jul 2023 11:29:04	U	SMPL
O3796-01MSD - 1	TOPSOIL-OU3-DENALI-TEMPMSD	17465	3.6638	-	7470A	PPB	31 Jul 2023 11:31:19	U	SMPL
CCV30 - 1	CCV30	23334	4.8892	-	7470A	PPB	31 Jul 2023 11:33:34	U	SMPL
CCB30 - 1	CCB30	-18	0.0135	-	7470A	PPB	31 Jul 2023 11:35:49	U	SMPL
PB154515BL - 1	PBW	69	0.0317	-	7470A	PPB	31 Jul 2023 11:38:04	U	SMPL
PB154515BS - 1	LCSW	17278	3.6248	-	7470A	PPB	31 Jul 2023 11:43:12	U	SMPL
O3793-01 - 1	D3156	13	0.0200	-	7470A	PPB	31 Jul 2023 11:45:26	U	SMPL
O3797-02 - 1	HANOVER-ST-1	68	0.0315	-	7470A	PPB	31 Jul 2023 11:47:42	U	SMPL
O3797-05 - 1	HANOVER-ST-2	54	0.0286	-	7470A	PPB	31 Jul 2023 11:49:58	U	SMPL
O3810-10 - 1	WC-01	59	0.0298	-	7470A	PPB	31 Jul 2023 11:52:14	U	SMPL
O3813-02 - 1	PLAUDERVILLE-1	75	0.0330	-	7470A	PPB	31 Jul 2023 11:54:29	U	SMPL
O3813-05 - 1	PLAUDERVILLE-2	55	0.0288	-	7470A	PPB	31 Jul 2023 11:56:46	U	SMPL
O3813-08 - 1	PLAUDERVILLE-3	80	0.0340	-	7470A	PPB	31 Jul 2023 11:59:03	U	SMPL
O3819-01 - 1	368	57	0.0292	-	7470A	PPB	31 Jul 2023 12:01:18	U	SMPL
CCV31 - 1	CCV31	22928	4.8044	-	7470A	PPB	31 Jul 2023 12:03:34	U	SMPL
CCB31 - 1	CCB31	-31	0.0108	-	7470A	PPB	31 Jul 2023 12:05:49	U	SMPL
O3819-01DUP - 1	368DUP	38	0.0252	-	7470A	PPB	31 Jul 2023 12:08:05	U	SMPL
O3819-01MS - 1	368MS	16152	3.3897	-	7470A	PPB	31 Jul 2023 12:10:21	U	SMPL
O3819-01MSD - 1	368MSD	16179	3.3953	-	7470A	PPB	31 Jul 2023 12:12:36	U	SMPL
PB154516BL - 1	PBW	-45	0.0079	-	7470A	PPB	31 Jul 2023 12:14:51	U	SMPL
PB154516BS - 1	LCSW	15723	3.3001	-	7470A	PPB	31 Jul 2023 12:17:07	U	SMPL
O3819-02 - 1	370	2	0.0177	-	7470A	PPB	31 Jul 2023 12:19:24	U	SMPL
O3819-02DUP - 1	370DUP	50	0.0277	-	7470A	PPB	31 Jul 2023 12:21:40	U	SMPL
O3819-02MS - 1	370MS	15958	3.3492	-	7470A	PPB	31 Jul 2023 12:23:56	U	SMPL
O3819-02MSD - 1	370MSD	16030	3.3642	-	7470A	PPB	31 Jul 2023 12:26:13	U	SMPL
CCV32 - 1	CCV32	22075	4.6263	-	7470A	PPB	31 Jul 2023 12:28:28	U	SMPL
CCB32 - 1	CCB32	-112	-0.0061	-	7470A	PPB	31 Jul 2023 12:30:43	U	SMPL
PB154514TB - 1	PB154486TB	-59	0.0050	-	7470A	PPB	31 Jul 2023 12:32:59	U	SMPL
PB154489TB - 1	PB154489TB	68	0.0315	-	7470A	PPB	31 Jul 2023 12:35:14	U	SMPL
PB154502TB - 1	PB154502TB	25	0.0225	-	7470A	PPB	31 Jul 2023 12:37:30	U	SMPL
O3796-01LX5 - 1		8	0.0190	-	7470A	PPB	31 Jul 2023 12:39:45	U	SMPL
O3796-01A - 1		18489	3.8776	-	7470A	PPB	31 Jul 2023 12:42:01	U	SMPL
O3819-01LX5 - 1		-41	0.0087	-	7470A	PPB	31 Jul 2023 12:44:17	U	SMPL
O3819-01A - 1		18953	3.9745	-	7470A	PPB	31 Jul 2023 12:46:33	U	SMPL
O3819-02LX5 - 1		-14	0.0144	-	7470A	PPB	31 Jul 2023 12:48:49	U	SMPL
O3819-02A - 1		19309	4.0488	-	7470A	PPB	31 Jul 2023 12:51:06	U	SMPL
CCV33 - 1	CCV33	23899	5.0071	-	7470A	PPB	31 Jul 2023 12:53:23	U	SMPL
CCB33 - 1	CCB33	-62	0.0044	-	7470A	PPB	31 Jul 2023 12:55:38	U	SMPL