

L312666.8


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

62560



A= 0.0000e+000

B= 1.5796e-004

C= 3.1147e-003

Rho= 0.9997017

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.008	0.008	29	0.000	29					0/0)
0.2	0.200	0.154	-0.046	957	0.0 %	957					-23
2.5	2.500	2.430	-0.070	15366	0.0 %	15366					133
5.0	5.000	5.154	0.154	32611	0.0 %	32611					3
7.5	7.500	7.569	0.069	47895	0.0 %	47895					1
10.0	10.000	9.885	-0.115	62560	0.0 %	62560					1

LB126668 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Jul 2023 15:13:44

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.0000	7471B	PPB	31 Jul 2023 15:15:37	S	Std
0.2 - 1	51.2	957	-	0.2000	7471B	PPB	31 Jul 2023 15:17:54	S	Std
2.5 - 1	52.5	15366	-	2.5000	7471B	PPB	31 Jul 2023 15:20:10	S	Std
5.0 - 1	55	32611	-	5.0000	7471B	PPB	31 Jul 2023 15:22:27	S	Std
7.5 - 1	57.5	47895	-	7.5000	7471B	PPB	31 Jul 2023 15:27:20	S	Std
10.0 - 1	511	62560	-	10.0000	7471B	PPB	31 Jul 2023 15:34:47	S	Std
ICV93 - 1	ICV93	24685	3.9023	-	7471B	PPB	31 Jul 2023 15:38:47	U	SMPL
ICB93 - 1	ICB93	-51	-0.0049	-	7471B	PPB	31 Jul 2023 15:41:01	U	SMPL
CCV34 - 1	CCV34	30026	4.7460	-	7471B	PPB	31 Jul 2023 15:43:16	U	SMPL
CCB34 - 1	CCB34	-57	-0.0059	-	7471B	PPB	31 Jul 2023 15:45:30	U	SMPL
CRA - 1	CRA	920	0.1484	-	7471B	PPB	31 Jul 2023 15:47:45	U	SMPL
HighStd - 1	HighStd	63051	9.9625	-	7471B	PPB	31 Jul 2023 15:52:34	U	SMPL
ChkStd - 1	ChkStd	44652	7.0562	-	7471B	PPB	31 Jul 2023 15:54:48	U	SMPL
PB154539BL - 1	PBS	-63	-0.0068	-	7471B	PPB	31 Jul 2023 15:57:05	U	SMPL
PB154539BS - 1	LCSS	20998	3.3199	-	7471B	PPB	31 Jul 2023 15:59:21	U	SMPL
O3598-01 - 1	RB-52-(0-2)	46563	7.3581	-	7471B	PPB	31 Jul 2023 16:01:37	U	SMPL
O3598-01DUP - 1	RB-52-(0-2)DUP	44817	7.0823	-	7471B	PPB	31 Jul 2023 16:03:53	U	SMPL
O3598-01MS - 1	RB-52-(0-2)MS	71706	11.3296	-	7471B	PPB	31 Jul 2023 16:06:24	U	SMPL
O3598-01MSD - 1	RB-52-(0-2)MSD	67756	10.7057	-	7471B	PPB	31 Jul 2023 16:08:42	U	SMPL
O3598-02 - 1	RB-52-(2-5)	4866	0.7717	-	7471B	PPB	31 Jul 2023 16:10:58	U	SMPL
CCV35 - 1	CCV35	29822	4.7137	-	7471B	PPB	31 Jul 2023 16:13:14	U	SMPL
CCB35 - 1	CCB35	-60	-0.0064	-	7471B	PPB	31 Jul 2023 16:15:29	U	SMPL
O3598-03 - 1	RB-54-(0-2)	7436	1.1777	-	7471B	PPB	31 Jul 2023 16:17:45	U	SMPL
O3598-04 - 1	RB-54-(2-5)	10522	1.6651	-	7471B	PPB	31 Jul 2023 16:20:00	U	SMPL
O3598-05 - 1	RB-55-(0-2)	14983	2.3698	-	7471B	PPB	31 Jul 2023 16:22:15	U	SMPL
O3598-06 - 1	RB-55-(2-5)	50855	8.0361	-	7471B	PPB	31 Jul 2023 16:24:30	U	SMPL
O3598-07 - 1	RB-56-(0-2)	42566	6.7267	-	7471B	PPB	31 Jul 2023 16:26:46	U	SMPL
O3598-08 - 1	RB-56-(2-5)	10603	1.6779	-	7471B	PPB	31 Jul 2023 16:29:03	U	SMPL
O3598-09 - 1	RB-53-(0-2)	60548	9.5671	-	7471B	PPB	31 Jul 2023 16:31:20	U	SMPL
O3598-10 - 1	RB-53-(2-5)	12547	1.9850	-	7471B	PPB	31 Jul 2023 16:33:36	U	SMPL
O3598-11 - 1	RB-57-(1-3)	5435	0.8616	-	7471B	PPB	31 Jul 2023 16:35:54	U	SMPL
O3598-12 - 1	RB-57-(5-7)	7637	1.2094	-	7471B	PPB	31 Jul 2023 16:38:10	U	SMPL
CCV36 - 1	CCV36	30005	4.7426	-	7471B	PPB	31 Jul 2023 16:40:25	U	SMPL
CCB36 - 1	CCB36	-23	-0.0005	-	7471B	PPB	31 Jul 2023 16:42:41	U	SMPL
PB154540BL - 1	PBS	33	0.0083	-	7471B	PPB	31 Jul 2023 16:44:57	U	SMPL
PB154540BS - 1	LCSS	22695	3.5880	-	7471B	PPB	31 Jul 2023 16:47:12	U	SMPL
O3815-01 - 1	BUILDING-A-1-(0-5)	55267	8.7330	-	7471B	PPB	31 Jul 2023 16:49:27	U	SMPL
O3815-02 - 1	BUILDING-A-2-(0-5)	35043	5.5384	-	7471B	PPB	31 Jul 2023 16:51:43	U	SMPL
O3815-03 - 1	DUP-01	32414	5.1232	-	7471B	PPB	31 Jul 2023 16:53:59	U	SMPL
O3815-04 - 1	BUILDING-B-21-(0-5)	1450	0.2322	-	7471B	PPB	31 Jul 2023 16:56:15	U	SMPL
O3815-05 - 1	BUILDING-B-7-(0-5)	4025	0.6389	-	7471B	PPB	31 Jul 2023 16:58:31	U	SMPL
O3821-01 - 1	S-1R	1008	0.1623	-	7471B	PPB	31 Jul 2023 17:00:48	U	SMPL
O3826-01 - 1	HR-01-072823	4754	0.7540	-	7471B	PPB	31 Jul 2023 17:03:04	U	SMPL
O3833-01 - 1	82ND-1	787	0.1274	-	7471B	PPB	31 Jul 2023 17:05:20	U	SMPL
CCV37 - 1	CCV37	29930	4.7308	-	7471B	PPB	31 Jul 2023 17:07:35	U	SMPL
CCB37 - 1	CCB37	-33	-0.0021	-	7471B	PPB	31 Jul 2023 17:09:51	U	SMPL
O3834-01 - 1	B-P2A	1584	0.2533	-	7471B	PPB	31 Jul 2023 17:12:06	U	SMPL
O3834-02 - 1	B-P2B	1504	0.2407	-	7471B	PPB	31 Jul 2023 17:14:21	U	SMPL
O3835-01 - 1	HR-01-072823	1836	0.2931	-	7471B	PPB	31 Jul 2023 17:16:36	U	SMPL
O3835-01DUP - 1	HR-01-072823DUP	1727	0.2759	-	7471B	PPB	31 Jul 2023 17:18:52	U	SMPL
O3835-01MS - 1	HR-01-072823MS	24920	3.9394	-	7471B	PPB	31 Jul 2023 17:21:08	U	SMPL
O3835-01MSD - 1	HR-01-072823MSD	27172	4.2951	-	7471B	PPB	31 Jul 2023 17:23:24	U	SMPL
O3598-01LX5 - 1		9748	1.5429	-	7471B	PPB	31 Jul 2023 17:25:40	U	SMPL
O3598-01A - 1		71086	11.2317	-	7471B	PPB	31 Jul 2023 17:27:57	U	SMPL
O3835-01LX5 - 1		278	0.0470	-	7471B	PPB	31 Jul 2023 17:30:14	U	SMPL
O3835-01A - 1		23520	3.7183	-	7471B	PPB	31 Jul 2023 17:32:30	U	SMPL
CCV38 - 1	CCV38	30749	4.8602	-	7471B	PPB	31 Jul 2023 17:34:46	U	SMPL
CCB38 - 1	CCB38	-37	-0.0027	-	7471B	PPB	31 Jul 2023 17:37:01	U	SMPL
O3598-01DLX2 - 1		22287	3.5235	-	7471B	PPB	31 Jul 2023 17:39:16	U	SMPL
O3598-01DUPDLX2 - 1		21887	3.4603	-	7471B	PPB	31 Jul 2023 17:41:31	U	SMPL
O3598-01MSDLX2 - 1		34232	5.4103	-	7471B	PPB	31 Jul 2023 17:43:47	U	SMPL
O3598-01MSDDLX2 - 1		35493	5.6095	-	7471B	PPB	31 Jul 2023 17:46:02	U	SMPL
O3598-01LDLX10 - 1		4523	0.7176	-	7471B	PPB	31 Jul 2023 17:48:17	U	SMPL
O3598-01ADLX2 - 1		36287	5.7349	-	7471B	PPB	31 Jul 2023 17:50:33	U	SMPL
CCV39 - 1	CCV39	29975	4.7379	-	7471B	PPB	31 Jul 2023 17:52:50	U	SMPL
CCB39 - 1	CCB39	-32	-0.0019	-	7471B	PPB	31 Jul 2023 17:55:06	U	SMPL