



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126669

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP103523
Residual chlorine ICV-LCS, 0.4PPM	WP103521
Chlorine Calibration std, 0.1ppm	WP103517
Chlorine Calibration std, 0.2ppm	WP103518
Chlorine Calibration std, 0.8ppm	WP103520
Chlorine Calibration std, 0.0ppm	WP103516
Chlorine Calibration std, 1.6ppm	WP103519
Residual Chlorine Calibration and CCV std, 0	WP103522
Total Chlorine Powder Pillows	W2759

Intercept: 0.0090

Slope: 1.0046

Regression: 0.999957

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		07/31/2023	10:30
2	CAL2	0.1	1	0.000	0.112	0.112	0.10	3	07/31/2023	10:33
3	CAL3	0.2	1	0.000	0.211	0.211	0.20	0.5	07/31/2023	10:36
4	CAL4	0.4	1	0.000	0.414	0.414	0.40	0.8	07/31/2023	10:39
5	CAL5	0.8	1	0.000	0.819	0.819	0.81	0.8	07/31/2023	10:42
6	CAL6	1.6	1	0.000	1.612	1.612	1.60	-0.3	07/31/2023	10:45

Analysis Method: SM4500 Cl GANALYST: RubinaParameter: Residual ChlorineSUPERVISOR REVIEW BY: SohilRun Number: LB126669

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4210	0.4210	0.4110	07/31/2023	10:48
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	07/31/2023	10:51
3	CCV1			0.4	1	0.0000	0.4120	0.4120	0.4020	07/31/2023	10:54
4	CCB1				1	0.0000	0.0010	0.0010	-0.0070	07/31/2023	10:57
5	LB126669BL	50	50		1	0.0000	0.0010	0.0010	-0.0070	07/31/2023	11:00
6	LB126669BS	50	50	0.4	1	0.0000	0.4020	0.4020	0.3920	07/31/2023	11:03
7	O3572-09	50	50		1	0.0000	0.0620	0.0620	0.0540	07/31/2023	11:06
8	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4000	07/31/2023	11:09
9	CCB2				1	0.0000	0.0000	0.0000	-0.0080	07/31/2023	11:12