



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126671

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/31/2023	15:02
2	CAL2	1	Water	NA	NA	20.7	7.00	07/31/2023	15:05
3	CAL3	1	Water	NA	NA	20.7	10.02	07/31/2023	15:10
4	ICV	1	Water	NA	NA	20.7	6.99	07/31/2023	15:15
5	CCV1	1	Water	NA	NA	20.7	2.01	07/31/2023	15:20
6	O3833-03	1	Solid	20.02	20	23.4	6.10	07/31/2023	15:25
7	O3849-02	1	Solid	20.01	20	20.5	8.90	07/31/2023	15:30
8	O3849-02DUP	1	Solid	20.02	20	20.6	8.91	07/31/2023	15:32
9	CCV2	1	Water	NA	NA	20.7	12.03	07/31/2023	15:35

WORKLIST(Hardcopy Internal Chain)

W 12667)

WorkList Name : CORROSIONITY O3849 WorkList ID : 172346 Department : Wet-Chemistry Date : 07-31-2023 15:22:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3833-03	82ND-1	Solid	Corrosivity	Cool 4 deg C	DIST01	L12	07/27/2023	9045D
O3849-02	GROVE-SOIL-0731	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	07/31/2023	9045D

Date/Time 07/31/2023 15:00
 Raw Sample Received by: JDCM
 Raw Sample Relinquished by: JDCM

Date/Time 07-31-2023 17:00
 Raw Sample Received by: JDCM
 Raw Sample Relinquished by: JDCM