



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: Sohil

Run Number: LB126672

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP101736

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB126672BL		1	22.4	0.81	0.85	07/31/2023	14:20
2	LB126672BS	1413	1	22.3	1332.00	1404.43	07/31/2023	14:25
3	O3824-01		1	22.6	5.02	5.26	07/31/2023	14:30
4	O3824-01DUP		1	22.6	5.03	5.27	07/31/2023	14:35

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

WORKLIST(Hardcopy Internal Chain)

L13126672

WorkList Name : CONDUCTANCE-07282

WorkList ID : 172307

Department : Wet-Chemistry

Date : 07-28-2023 12:45:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3824-01	JS9600	Water	Conductance	Cool 4 deg C	STER17	L11	06/30/2023	SM2510 B

Date/Time 07/31/23 14:50
 Raw Sample Received by: 12/25/21
 Raw Sample Relinquished by: 12/25/21

Date/Time 07/31/23 14:40
 Raw Sample Received by: 12/25/21
 Raw Sample Relinquished by: 12/25/21