



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126683

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	08/01/2023	08:40
2	CAL2	1	Water	NA	NA	20.7	7.00	08/01/2023	08:41
3	CAL3	1	Water	NA	NA	20.7	10.03	08/01/2023	08:45
4	ICV	1	Water	NA	NA	20.7	7.01	08/01/2023	08:47
5	CCV1	1	Water	NA	NA	20.8	2.01	08/01/2023	08:50
6	O3824-01	1	Water	NA	NA	22.6	4.47	08/01/2023	09:00
7	O3843-01	1	Water	NA	NA	23.5	5.86	08/01/2023	09:05
8	O3843-01DUP	1	Water	NA	NA	23.6	5.87	08/01/2023	09:06
9	CCV2	1	Water	NA	NA	20.7	12.02	08/01/2023	09:10

VB 126683

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 03824 WorkList ID : 172367 Department : Wet-Chemistry Date : 08-01-2023 08:24:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3824-01	JS9600	Water	pH	Cool 4 deg C	STER17	L11	06/30/2023	9040C
O3843-01	NWB-1935	Water	pH	Cool 4 deg C	PSEG03	L11	07/31/2023	9040C

Date/Time 08/01/23 08:30
Raw Sample Received by: JD (CSM)
Raw Sample Relinquished by: JD (CSM)

Date/Time 08/01/23 11:00
Raw Sample Received by: JD (CSM)
Raw Sample Relinquished by: JD (CSM)